

MEDICAL LABORATORY EVALUATION

PARTICIPANT SUMMARY

2 • 0 • 1 • 9

Hematology, Coagulation,
Blood Bank, Urinalysis, PPM
2019 MLE-M2

**ACP | Medical Laboratory
Evaluation** 

Total Commitment to Education and Service
Provided by ACP, Inc.

Table of Contents

Evaluation Criteria	5
Hematology	
HemoCue (HQ Samples – Module 215).....	6
Hemoglobin.....	6
Glucose.....	6
Sedimentation Rate.....	6
Hematology with 5-part Automated Differential (CL Samples – Module 223).....	7
White Blood Cell Count.....	7
Red Blood Cell Count.....	7
Hemoglobin.....	8
Hematocrit.....	8
Platelet Count.....	9
Automated Differential.....	9
Sysmex Hematology with 3-part Automated Differential (SYX Samples – Module 224).....	12
White Blood Cell Count.....	12
Red Blood Cell Count.....	13
Hemoglobin.....	14
Hematocrit.....	15
Platelet Count.....	16
Automated Differential.....	17
Basic Hematology with 3-part Automated Differential (HD Samples – Module 225).....	20
White Blood Cell Count.....	20
Red Blood Cell Count.....	21
Hemoglobin.....	23
Hematocrit.....	24
Platelet Count.....	26
Automated Differential.....	27
Hematology with 5-part Automated Differential (DIF Samples – Module 226).....	32
White Blood Cell Count.....	32
Red Blood Cell Count.....	32
Hemoglobin.....	33
Hematocrit.....	33
Platelet Count.....	34
Automated Differential.....	34
Blood Lead.....	37
Reticulocyte Count.....	37

Table of Contents (cont'd)

Hematology (cont'd)

Hematology with 5-part Automated Differential (BCX Samples – Module 228)	38
White Blood Cell Count	38
Red Blood Cell Count	39
Hemoglobin	40
Hematocrit	41
Platelet Count	42
Automated Differential	43
Hematology with 5 or 6-part Automated Differential (MX Samples – Module 229)	48
White Blood Cell Count	48
Red Blood Cell Count	49
Hemoglobin	50
Hematocrit	50
Platelet Count	51
Automated Differential (including Immature Granulocytes)	52
Waived Hematology (HD Samples – Module 213)	75
Hemoglobin	75
Hematocrit	75
Blood Cell Identification	59

Blood Bank

ABO Group	64
Rh Factor (D Type)	64
Unexpected Antibody Detection	64
Antibody Identification	65
Compatibility Testing	65

Coagulation

Prothrombin Time	66
International Normalized Ratio (INR)	67
Activated Partial Thromboplastin Time	68
Fibrinogen	68
CoaguChek XS Plus Prothrombin Time	69
International Normalized Ratio (INR)	70
CoaguChek XS INR	70
i-Stat Prothrombin Time	71
International Normalized Ratio (INR)	71
Fluid Cell Count/Crystals	71
WBC Count	71
RBC Count	71
Crystal Identification	72

Table of Contents (cont'd)

Urinalysis

Microalbumin, Urine	74
Dipstick	74
Quantitative	74
Creatinine, Urine	74
Dipstick	74
Quantitative	74
Urinalysis Dipstick	76
Specific Gravity	76
pH	77
Protein.....	78
Glucose.....	80
Ketones.....	81
Bilirubin	82
Urobilinogen.....	83
Blood or Hemoglobin.....	84
Leukocyte Esterase.....	85
Nitrite.....	86
Microalbumin (Dipstick Only).....	87
Urine hCG	88
Fecal Occult Blood	89
KOH Skin Preparation	75
Urine Sediment Identification	90

Provider-Performed Microscopy (PPM)

Provider-Performed Microscopy (PPM)	92
--	----

EVALUATION CRITERIA

The evaluation criteria used in the MLE Program is in accordance with the Clinical Laboratory Improvement Amendments of 1988 (CLIA '88) federal requirements for proficiency testing. The criteria are included below.

Qualitative

For qualitative procedures, evaluation is based on participant or referee consensus. If participant consensus is not reached, CMS requirements call for grading by referee consensus. A minimum percentage of participants or referee laboratories must receive a passing score or the challenge is not evaluated due to lack of consensus. These percentages are listed below.

ABO Group	95% Participant or 100% Referee Consensus
Antibody Identification	95% Consensus
Blood Cell Identification	80% Consensus
Compatibility Testing	95% Participant or 100% Referee Consensus
Creatinine (Semi-Quantitative)	80% Consensus
Crystal Identification	80% Consensus
Fecal Occult Blood	80% Consensus
KOH Skin Preparation	80% Consensus
Microalbumin (Semi-Quantitative)	80% Consensus
Provider-Performed Microscopy	80% Consensus
Rh Factor (D Type)	95% Participant or 100% Referee Consensus
Unexpected Antibody Detection	95% Consensus
Urine Dipstick	80% Consensus
Urine hCG	80% Consensus
Urine Sediment Identification	80% Consensus

Quantitative

For quantitative procedures, a mean and standard deviation (SD) are calculated for each peer group consisting of 10 or more laboratories except for Coagulation (CG Specimens) which consist of peer groups of 5 or more laboratories. Acceptable performance is established on a target value \pm the intervals below. An explanation on how to calculate the range of acceptability based upon these limits is also provided in your MLE Program Guide on page 37 under the heading "Acceptable Ranges for Quantitative Results."

Activated Partial Thromboplastin Time	$\pm 15\%$
Automated Differential	± 3 SD
Blood Lead	$\pm 4 \mu\text{g/dL}$ or $\pm 10\%^*$
Body Fluid - Red Cell Count	± 2 SD
Body Fluid - White Cell Count	± 2 SD
Creatinine, Urine (Quantitative)	$\pm 17\%$
Fibrinogen	$\pm 20\%$
Hematocrit	$\pm 6\%$
Hematocrit, Waived	$\pm 6\%$ or ± 2 SD*
Hemoglobin	$\pm 7\%$
Hemoglobin, Waived	$\pm 7\%$ or ± 2 SD*
International Normalized Ratio (INR)	$\pm 20\%$
Microalbumin (Quantitative)	$\pm 30\%$
Platelet Count	$\pm 25\%$
Prothrombin Time	$\pm 15\%$
Red Blood Cell Count	$\pm 6\%$
Reticulocyte Count	$\pm 30\%$ or ± 2 SD*
Sedimentation Rate	± 2 SD
Specific Gravity	± 0.010
White Blood Cell Count	$\pm 15\%$
Whole Blood Glucose – HemoCue	$\pm 12 \text{ mg/dL}$ or $\pm 20\%^*$

*Whichever is greater

HEMOCUE HEMATOLOGY–HEMOGLOBIN (g/dL)

<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>Specimen HQ-3</u>				<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>Specimen HQ-4</u>			
			<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>				<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	32	13.12	0.27	2.1	13.1	12.2 - 14.1	34	4.87	0.23	4.8	4.9	4.5 - 5.3	
HemoCue 201/+	32	13.12	0.27	2.1	13.1	12.2 - 14.1	34	4.87	0.23	4.8	4.9	4.5 - 5.3	

HEMOCUE HEMATOLOGY–GLUCOSE (mg/dL)

<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>Specimen HQ-3</u>				<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>Specimen HQ-4</u>			
			<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>				<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	30	118.2	7.3	6.2	117	94 - 142	29	43.0	5.0	11.7	43	30 - 55	
All HemoCue Methods	28	118.6	7.4	6.3	117	94 - 143	27	43.4	4.9	11.4	43	31 - 56	
HemoCue Glucose 201/+	27	118.3	7.4	6.2	117	94 - 142	27	43.4	4.9	11.4	43	31 - 56	

SEDIMENTATION RATE (MM/HR)

<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>Specimen ES-3</u>				<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>Specimen ES-4</u>			
			<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>				<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	95	44.0	10.5	23.7	41	23 - 65	96	7.1	2.8	39.1	7	1 - 13	
All Automated Methods	25	55.2	9.9	17.9	54	35 - 75	25	7.8	3.4	43.8	7	0 - 15	
All Manual Methods	70	40.6	8.6	21.3	38	23 - 58	71	7.1	2.7	38.5	7	1 - 13	
All Vital Diagnostics Methods	14	54.3	8.4	15.5	52	37 - 72	14	6.7	1.8	26.4	7	3 - 11	
Westergren - diluted	59	40.5	8.1	19.9	38	24 - 57	60	6.8	2.6	38.5	7	1 - 13	

SEDIMAT SEDIMENTATION RATE (MM/HR)

<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>Specimen MAT-3</u>				<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>Specimen MAT-4</u>			
			<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>				<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
Polymedco Sedimat 15	12	63.5	6.2	9.8	63	51 - 76	12	2.5	0.9	36.2	3	0 - 5	

HEMATOLOGY W/ 5-PART DIFFERENTIAL–WHITE BLOOD CELL COUNT (x K/uL)

<u>Instrument</u>	Specimen CL-6						Specimen CL-7					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	14	7.21	0.49	6.8	7.3	6.1 - 8.3	14	9.48	0.53	5.6	9.6	8.0 - 11.0
All Abbott Cell-Dyn Instruments	12	7.27	0.10	1.3	7.3	6.1 - 8.4	12	9.57	0.21	2.2	9.6	8.1 - 11.1
Abbott Cell-Dyn Ruby	10	7.27	0.10	1.3	7.3	6.1 - 8.4	10	9.57	0.21	2.2	9.6	8.1 - 11.1
	Specimen CL-8						Specimen CL-9					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	14	2.93	0.25	8.5	3.0	2.4 - 3.4	14	19.44	1.23	6.3	19.6	16.5 - 22.4
All Abbott Cell-Dyn Instruments	12	2.99	0.13	4.5	3.0	2.5 - 3.5	12	19.49	0.49	2.5	19.5	16.5 - 22.5
Abbott Cell-Dyn Ruby	10	2.99	0.13	4.5	3.0	2.5 - 3.5	10	19.49	0.49	2.5	19.5	16.5 - 22.5
	Specimen CL-10											
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	14	2.99	0.32	10.7	3.0	2.5 - 3.5						
All Abbott Cell-Dyn Instruments	12	3.07	0.20	6.4	3.0	2.6 - 3.6						
Abbott Cell-Dyn Ruby	10	3.07	0.20	6.4	3.0	2.6 - 3.6						

HEMATOLOGY W/ 5-PART DIFFERENTIAL–RED BLOOD CELL COUNT (x M/uL)

<u>Instrument</u>	Specimen CL-6						Specimen CL-7					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	14	4.531	0.112	2.5	4.56	4.25 - 4.81	14	6.514	0.192	3.0	6.55	6.12 - 6.91
All Abbott Cell-Dyn Instruments	12	4.589	0.055	1.2	4.60	4.31 - 4.87	12	6.613	0.126	1.9	6.60	6.21 - 7.01
Abbott Cell-Dyn Ruby	10	4.589	0.055	1.2	4.60	4.31 - 4.87	10	6.613	0.126	1.9	6.60	6.21 - 7.01
	Specimen CL-8						Specimen CL-9					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	14	2.278	0.050	2.2	2.27	2.14 - 2.42	14	5.286	0.116	2.2	5.31	4.96 - 5.61
All Abbott Cell-Dyn Instruments	12	2.290	0.051	2.2	2.27	2.15 - 2.43	12	5.340	0.072	1.3	5.35	5.01 - 5.67
Abbott Cell-Dyn Ruby	10	2.290	0.051	2.2	2.27	2.15 - 2.43	10	5.340	0.072	1.3	5.35	5.01 - 5.67
	Specimen CL-10											
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	14	2.290	0.057	2.5	2.30	2.15 - 2.43						
All Abbott Cell-Dyn Instruments	12	2.306	0.035	1.5	2.30	2.16 - 2.45						
Abbott Cell-Dyn Ruby	10	2.306	0.035	1.5	2.30	2.16 - 2.45						

HEMATOLOGY W/ 5-PART DIFFERENTIAL–HEMOGLOBIN (g/dL)

<u>Instrument</u>	Specimen CL-6						Specimen CL-7					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	14	13.18	0.50	3.8	13.4	12.2 - 14.2	14	17.77	0.72	4.0	18.0	16.5 - 19.1
All Abbott Cell-Dyn Instruments	12	13.29	0.34	2.5	13.3	12.3 - 14.3	12	17.99	0.35	2.0	18.0	16.7 - 19.3
Abbott Cell-Dyn Ruby	10	13.29	0.34	2.5	13.3	12.3 - 14.3	10	17.99	0.35	2.0	18.0	16.7 - 19.3
<u>Instrument</u>	Specimen CL-8						Specimen CL-9					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	13	5.70	0.15	2.6	5.7	5.3 - 6.1	14	16.22	0.55	3.4	16.3	15.0 - 17.4
All Abbott Cell-Dyn Instruments	12	5.66	0.14	2.5	5.7	5.2 - 6.1	12	16.26	0.56	3.4	16.2	15.1 - 17.4
Abbott Cell-Dyn Ruby	10	5.66	0.14	2.5	5.7	5.2 - 6.1	10	16.26	0.56	3.4	16.2	15.1 - 17.4
<u>Instrument</u>	Specimen CL-10											
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>						
All Method	13	5.68	0.17	3.0	5.7	5.2 - 6.1						
All Abbott Cell-Dyn Instruments	12	5.63	0.16	2.8	5.6	5.2 - 6.1						
Abbott Cell-Dyn Ruby	10	5.63	0.16	2.8	5.6	5.2 - 6.1						

HEMATOLOGY W/ 5-PART DIFFERENTIAL–HEMATOCRIT (percent)

<u>Instrument</u>	Specimen CL-6						Specimen CL-7					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	14	38.49	1.71	4.5	37.8	36.1 - 40.8	14	53.92	1.88	3.5	53.5	50.6 - 57.2
All Abbott Cell-Dyn Instruments	12	37.63	0.55	1.5	37.8	35.3 - 39.9	12	53.09	1.03	1.9	53.1	49.9 - 56.3
Abbott Cell-Dyn Ruby	10	37.63	0.55	1.5	37.8	35.3 - 39.9	10	53.09	1.03	1.9	53.1	49.9 - 56.3
<u>Instrument</u>	Specimen CL-8						Specimen CL-9					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	14	16.33	0.97	5.9	16.0	15.3 - 17.4	14	46.85	2.55	5.4	45.8	44.0 - 49.7
All Abbott Cell-Dyn Instruments	12	15.81	0.23	1.4	15.8	14.8 - 16.8	12	45.50	0.77	1.7	45.4	42.7 - 48.3
Abbott Cell-Dyn Ruby	10	15.81	0.23	1.4	15.8	14.8 - 16.8	10	45.50	0.77	1.7	45.4	42.7 - 48.3
<u>Instrument</u>	Specimen CL-10											
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>						
All Method	14	16.45	1.07	6.5	16.0	15.4 - 17.5						
All Abbott Cell-Dyn Instruments	12	15.94	0.17	1.1	15.9	14.9 - 16.9						
Abbott Cell-Dyn Ruby	10	15.94	0.17	1.1	15.9	14.9 - 16.9						

HEMATOLOGY W/ 5-PART DIFFERENTIAL-PLATELET COUNT (x K/uL)

<u>Instrument</u>	Specimen CL-6						Specimen CL-7					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	14	263.0	8.7	3.3	264	197 - 329	13	129.3	30.9	23.9	143	96 - 162
All Abbott Cell-Dyn Instruments	12	264.4	9.1	3.5	267	198 - 331	12	148.0	15.2	10.3	148	111 - 185
Abbott Cell-Dyn Ruby	10	264.4	9.1	3.5	267	198 - 331	10	148.0	15.2	10.3	148	111 - 185
Specimen CL-8												
All Method	14	76.9	8.2	10.7	75	57 - 97	14	497.9	43.3	8.7	487	373 - 623
All Abbott Cell-Dyn Instruments	12	73.7	2.9	3.9	74	55 - 93	12	515.4	39.1	7.6	534	386 - 645
Abbott Cell-Dyn Ruby	10	73.7	2.9	3.9	74	55 - 93	10	515.4	39.1	7.6	534	386 - 645
Specimen CL-10												
All Method	14	78.2	13.2	16.8	75	58 - 98						
All Abbott Cell-Dyn Instruments	12	71.9	5.7	8.0	72	53 - 90						
Abbott Cell-Dyn Ruby	10	71.9	5.7	8.0	72	53 - 90						

HEMATOLOGY W/ 5-PART DIFFERENTIAL-NEUTROPHILS (percent)

<u>Instrument</u>	Specimen CL-6						Specimen CL-7					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	14	66.38	1.38	2.1	66.6	62.2 - 70.6	14	69.11	2.28	3.3	68.8	62.2 - 76.0
All Abbott Cell-Dyn Instruments	12	66.39	1.31	2.0	66.4	62.4 - 70.4	12	68.51	0.69	1.0	68.5	66.4 - 70.6
Abbott Cell-Dyn Ruby	10	66.39	1.31	2.0	66.4	62.4 - 70.4	10	68.51	0.69	1.0	68.5	66.4 - 70.6
Specimen CL-8												
All Method	14	45.79	2.00	4.4	45.8	39.7 - 51.8	14	73.97	1.34	1.8	74.2	69.9 - 78.0
All Abbott Cell-Dyn Instruments	12	45.36	0.88	1.9	45.7	42.7 - 48.0	12	73.81	0.72	1.0	74.1	71.6 - 76.0
Abbott Cell-Dyn Ruby	10	45.36	0.88	1.9	45.7	42.7 - 48.0	10	73.81	0.72	1.0	74.1	71.6 - 76.0
Specimen CL-10												
All Method	14	46.92	2.17	4.6	47.0	40.4 - 53.5						
All Abbott Cell-Dyn Instruments	12	47.01	1.61	3.4	46.8	42.1 - 51.9						
Abbott Cell-Dyn Ruby	10	47.01	1.61	3.4	46.8	42.1 - 51.9						

HEMATOLOGY W/ 5-PART DIFFERENTIAL–LYMPHOCYTES (percent)

<u>Instrument</u>	Specimen CL-6						Specimen CL-7					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	14	23.88	4.52	18.9	24.8	10.3 - 37.5	14	19.73	3.88	19.6	21.0	8.1 - 31.4
All Abbott Cell-Dyn Instruments	12	25.47	1.40	5.5	25.8	21.2 - 29.7	12	21.19	0.33	1.6	21.1	20.1 - 22.2
Abbott Cell-Dyn Ruby	10	25.47	1.40	5.5	25.8	21.2 - 29.7	10	21.19	0.33	1.6	21.1	20.1 - 22.2
<u>Instrument</u>	Specimen CL-8						Specimen CL-9					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	14	39.29	6.56	16.7	41.8	19.5 - 59.0	14	16.84	2.97	17.6	17.7	7.9 - 25.8
All Abbott Cell-Dyn Instruments	12	41.24	4.04	9.8	42.6	29.1 - 53.4	12	17.63	0.77	4.4	17.5	15.3 - 20.0
Abbott Cell-Dyn Ruby	10	41.24	4.04	9.8	42.6	29.1 - 53.4	10	17.63	0.77	4.4	17.5	15.3 - 20.0
<u>Instrument</u>	Specimen CL-10											
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>						
All Method	14	39.32	6.18	15.7	40.9	20.7 - 57.9						
All Abbott Cell-Dyn Instruments	12	41.97	1.48	3.5	42.3	37.5 - 46.5						
Abbott Cell-Dyn Ruby	10	41.97	1.48	3.5	42.3	37.5 - 46.5						

HEMATOLOGY W/ 5-PART DIFFERENTIAL–MONOCYTES (percent)

<u>Instrument</u>	Specimen CL-6						Specimen CL-7					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	14	6.42	4.40	68.5	5.0	0.0 - 19.7	14	6.22	3.98	63.9	5.1	0.0 - 18.2
All Abbott Cell-Dyn Instruments	12	5.03	0.39	7.8	5.0	3.8 - 6.3	12	4.91	0.36	7.4	4.9	3.8 - 6.1
Abbott Cell-Dyn Ruby	10	5.03	0.39	7.8	5.0	3.8 - 6.3	10	4.91	0.36	7.4	4.9	3.8 - 6.1
<u>Instrument</u>	Specimen CL-8						Specimen CL-9					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	14	10.51	6.43	61.2	7.4	0.0 - 29.9	14	4.75	3.43	72.3	3.8	0.0 - 15.1
All Abbott Cell-Dyn Instruments	12	8.97	3.34	37.2	7.3	0.0 - 19.0	12	3.74	0.24	6.3	3.8	3.0 - 4.5
Abbott Cell-Dyn Ruby	10	8.97	3.34	37.2	7.3	0.0 - 19.0	10	3.74	0.24	6.3	3.8	3.0 - 4.5
<u>Instrument</u>	Specimen CL-10											
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>						
All Method	14	9.46	6.72	71.1	7.5	0.0 - 29.7						
All Abbott Cell-Dyn Instruments	12	6.96	0.71	10.2	7.2	4.8 - 9.1						
Abbott Cell-Dyn Ruby	10	6.96	0.71	10.2	7.2	4.8 - 9.1						

HEMATOLOGY W/ 5-PART DIFFERENTIAL–EOSINOPHILS (percent)

<u><i>Instrument</i></u>	Specimen CL-6						Specimen CL-7					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	14	2.95	0.84	28.6	2.9	0.4 - 5.5	14	4.68	2.09	44.6	5.3	0.0 - 11.0
All Abbott Cell-Dyn Instruments	12	2.89	0.13	4.7	2.9	2.4 - 3.3	12	5.23	0.23	4.4	5.3	4.5 - 6.0
Abbott Cell-Dyn Ruby	10	2.89	0.13	4.7	2.9	2.4 - 3.3	10	5.23	0.23	4.4	5.3	4.5 - 6.0
Specimen CL-8						Specimen CL-9						
All Method	14	3.91	1.47	37.6	4.2	0.0 - 8.4	14	4.14	1.15	27.8	4.6	0.6 - 7.6
All Abbott Cell-Dyn Instruments	12	4.17	0.29	7.0	4.2	3.2 - 5.1	12	4.57	0.29	6.4	4.6	3.6 - 5.5
Abbott Cell-Dyn Ruby	10	4.17	0.29	7.0	4.2	3.2 - 5.1	10	4.57	0.29	6.4	4.6	3.6 - 5.5
Specimen CL-10												
All Method	14	3.77	1.17	31.1	4.0	0.2 - 7.3						
All Abbott Cell-Dyn Instruments	12	3.89	0.43	11.0	4.0	2.6 - 5.2						
Abbott Cell-Dyn Ruby	10	3.89	0.43	11.0	4.0	2.6 - 5.2						

HEMATOLOGY W/ 5-PART DIFFERENTIAL–BASOPHILS (percent)

<u><i>Instrument</i></u>	Specimen CL-6						Specimen CL-7					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	14	0.37	0.32	85.5	0.2	0.0 - 1.4	14	0.25	0.25	100.2	0.2	0.0 - 1.1
All Abbott Cell-Dyn Instruments	12	0.23	0.10	41.6	0.2	0.0 - 0.6	12	0.14	0.10	68.3	0.1	0.0 - 0.5
Abbott Cell-Dyn Ruby	10	0.23	0.10	41.6	0.2	0.0 - 0.6	10	0.14	0.10	68.3	0.1	0.0 - 0.5
Specimen CL-8						Specimen CL-9						
All Method	14	0.49	0.58	118.0	0.2	0.0 - 2.3	14	0.28	0.23	82.1	0.2	0.0 - 1.0
All Abbott Cell-Dyn Instruments	12	0.26	0.31	120.6	0.1	0.0 - 1.2	12	0.21	0.15	68.3	0.2	0.0 - 0.7
Abbott Cell-Dyn Ruby	10	0.26	0.31	120.6	0.1	0.0 - 1.2	10	0.21	0.15	68.3	0.2	0.0 - 0.7
Specimen CL-10												
All Method	14	0.42	0.81	192.4	0.0	0.0 - 2.9						
All Abbott Cell-Dyn Instruments	12	0.01	0.04	264.6	0.0	0.0 - 0.2						
Abbott Cell-Dyn Ruby	10	0.01	0.04	264.6	0.0	0.0 - 0.2						

BASIC HEMATOLOGY W/ 3-PART DIFFERENTIAL–WHITE BLOOD CELL COUNT (x 10⁹/L) cont'd

Specimen HD-10

<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	437	1.97	0.15	7.8	2.0	1.6 - 2.3
All Abbott Cell-Dyn Instruments	109	2.04	0.13	6.2	2.0	1.7 - 2.4
All ABX Instruments	63	1.93	0.08	4.1	1.9	1.6 - 2.3
All Boule (CDS) Instruments	129	1.82	0.07	4.1	1.8	1.5 - 2.1
All COULTER Instruments	119	2.11	0.11	5.2	2.1	1.7 - 2.5
Abbott Cell-Dyn 1700	14	2.17	0.10	4.6	2.2	1.8 - 2.5
Abbott Cell-Dyn 1800	24	1.98	0.12	6.2	2.0	1.6 - 2.3
Abbott Cell-Dyn Emerald	70	2.04	0.11	5.2	2.0	1.7 - 2.4
Boule (CDS) Medonic M series	126	1.82	0.07	3.9	1.8	1.5 - 2.1
COULTER AcT diff/diff 2	114	2.11	0.11	5.1	2.1	1.7 - 2.5
Horiba ABX Micros/45/60	63	1.93	0.08	4.1	1.9	1.6 - 2.3

BASIC HEMATOLOGY W/ 3-PART DIFFERENTIAL–RED BLOOD CELL COUNT (x 10¹²/L)

Specimen HD-6

<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	434	4.636	0.108	2.3	4.64	4.35 - 4.92
All Abbott Cell-Dyn Instruments	106	4.582	0.122	2.7	4.58	4.30 - 4.86
All ABX Instruments	63	4.612	0.097	2.1	4.61	4.33 - 4.89
All Boule (CDS) Instruments	127	4.666	0.068	1.5	4.68	4.38 - 4.95
All COULTER Instruments	123	4.660	0.116	2.5	4.66	4.38 - 4.94
Abbott Cell-Dyn 1700	13	4.682	0.099	2.1	4.70	4.40 - 4.97
Abbott Cell-Dyn 1800	24	4.614	0.106	2.3	4.60	4.33 - 4.90
Abbott Cell-Dyn Emerald	70	4.547	0.129	2.8	4.56	4.27 - 4.82
Boule (CDS) Medonic M series	122	4.663	0.067	1.4	4.67	4.38 - 4.95
COULTER AcT diff/diff 2	118	4.657	0.118	2.5	4.65	4.37 - 4.94
Horiba ABX Micros/45/60	63	4.612	0.097	2.1	4.61	4.33 - 4.89

Specimen HD-7

<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
432	6.191	0.150	2.4	6.20	5.81 - 6.57
105	6.054	0.166	2.7	6.04	5.69 - 6.42
62	6.195	0.119	1.9	6.21	5.82 - 6.57
128	6.254	0.100	1.6	6.26	5.87 - 6.63
122	6.232	0.123	2.0	6.24	5.85 - 6.61
12	6.253	0.162	2.6	6.27	5.87 - 6.63
24	6.040	0.127	2.1	6.04	5.67 - 6.41
69	6.025	0.156	2.6	6.02	5.66 - 6.39
123	6.252	0.100	1.6	6.26	5.87 - 6.63
117	6.230	0.125	2.0	6.23	5.85 - 6.61
62	6.195	0.119	1.9	6.21	5.82 - 6.57

BASIC HEMATOLOGY W/ 3-PART DIFFERENTIAL-RED BLOOD CELL COUNT (x 10¹²/L) cont'd

<u>Instrument</u>	Specimen HD-8						Specimen HD-9					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	434	2.327	0.063	2.7	2.32	2.18 - 2.47	434	5.635	0.142	2.5	5.64	5.29 - 5.98
All Abbott Cell-Dyn Instruments	108	2.339	0.081	3.5	2.33	2.19 - 2.48	109	5.500	0.147	2.7	5.50	5.17 - 5.84
All ABX Instruments	62	2.285	0.045	2.0	2.29	2.14 - 2.43	63	5.617	0.099	1.8	5.63	5.27 - 5.96
All Boule (CDS) Instruments	131	2.303	0.043	1.9	2.30	2.16 - 2.45	129	5.738	0.096	1.7	5.74	5.39 - 6.09
All COULTER Instruments	121	2.358	0.057	2.4	2.36	2.21 - 2.50	121	5.646	0.133	2.3	5.63	5.30 - 5.99
Abbott Cell-Dyn 1700	14	2.384	0.070	3.0	2.40	2.24 - 2.53	14	5.569	0.158	2.8	5.63	5.23 - 5.91
Abbott Cell-Dyn 1800	23	2.416	0.063	2.6	2.43	2.27 - 2.57	23	5.490	0.133	2.4	5.47	5.16 - 5.82
Abbott Cell-Dyn Emerald	69	2.312	0.056	2.4	2.31	2.17 - 2.46	72	5.490	0.148	2.7	5.50	5.16 - 5.82
Boule (CDS) Medonic M series	126	2.301	0.042	1.8	2.30	2.16 - 2.44	124	5.739	0.097	1.7	5.74	5.39 - 6.09
COULTER AcT diff/diff 2	116	2.360	0.058	2.4	2.36	2.21 - 2.51	116	5.645	0.135	2.4	5.63	5.30 - 5.99
Horiba ABX Micros/45/60	62	2.285	0.045	2.0	2.29	2.14 - 2.43	63	5.617	0.099	1.8	5.63	5.27 - 5.96
Specimen HD-10												
All Method	432	2.326	0.062	2.7	2.32	2.18 - 2.47						
All Abbott Cell-Dyn Instruments	109	2.341	0.083	3.5	2.33	2.20 - 2.49						
All ABX Instruments	61	2.283	0.044	1.9	2.28	2.14 - 2.42						
All Boule (CDS) Instruments	130	2.306	0.039	1.7	2.31	2.16 - 2.45						
All COULTER Instruments	121	2.357	0.058	2.5	2.36	2.21 - 2.50						
Abbott Cell-Dyn 1700	14	2.397	0.053	2.2	2.40	2.25 - 2.55						
Abbott Cell-Dyn 1800	24	2.411	0.059	2.4	2.42	2.26 - 2.56						
Abbott Cell-Dyn Emerald	69	2.299	0.059	2.6	2.30	2.16 - 2.44						
Boule (CDS) Medonic M series	125	2.304	0.038	1.7	2.30	2.16 - 2.45						
COULTER AcT diff/diff 2	116	2.358	0.058	2.5	2.36	2.21 - 2.50						
Horiba ABX Micros/45/60	61	2.283	0.044	1.9	2.28	2.14 - 2.42						

BASIC HEMATOLOGY W/ 3-PART DIFFERENTIAL–HEMOGLOBIN (g/dL)

<u>Instrument</u>	Specimen HD-6						Specimen HD-7					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	436	13.43	0.25	1.8	13.4	12.4 - 14.4	435	17.04	0.30	1.8	17.0	15.8 - 18.3
All Abbott Cell-Dyn Instruments	107	13.41	0.25	1.9	13.4	12.4 - 14.4	106	17.02	0.31	1.8	17.0	15.8 - 18.3
All ABX Instruments	63	13.50	0.21	1.6	13.5	12.5 - 14.5	62	17.11	0.26	1.5	17.1	15.9 - 18.4
All Boule (CDS) Instruments	128	13.48	0.20	1.4	13.5	12.5 - 14.5	127	17.04	0.25	1.5	17.0	15.8 - 18.3
All COULTER Instruments	120	13.33	0.26	2.0	13.3	12.3 - 14.3	120	17.03	0.32	1.9	17.0	15.8 - 18.3
Abbott Cell-Dyn 1700	13	13.43	0.20	1.5	13.4	12.4 - 14.4	14	16.64	1.03	6.2	17.0	15.4 - 17.9
Abbott Cell-Dyn 1800	24	13.58	0.29	2.2	13.6	12.6 - 14.6	24	17.30	0.36	2.1	17.2	16.0 - 18.6
Abbott Cell-Dyn Emerald	70	13.35	0.23	1.7	13.3	12.4 - 14.3	71	16.94	0.29	1.7	16.9	15.7 - 18.2
Boule (CDS) Medonic M series	123	13.48	0.20	1.5	13.5	12.5 - 14.5	122	17.04	0.25	1.5	17.0	15.8 - 18.3
COULTER AcT diff/diff 2	115	13.32	0.27	2.0	13.3	12.3 - 14.3	115	17.03	0.32	1.9	17.0	15.8 - 18.3
Horiba ABX Micros/45/60	63	13.50	0.21	1.6	13.5	12.5 - 14.5	62	17.11	0.26	1.5	17.1	15.9 - 18.4

<u>Instrument</u>	Specimen HD-8						Specimen HD-9					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	436	5.87	0.14	2.4	5.9	5.4 - 6.3	436	18.21	0.34	1.9	18.2	16.9 - 19.5
All Abbott Cell-Dyn Instruments	109	5.87	0.19	3.3	5.9	5.4 - 6.3	109	18.15	0.36	2.0	18.1	16.8 - 19.5
All ABX Instruments	61	5.88	0.11	1.8	5.9	5.4 - 6.3	63	18.17	0.25	1.4	18.2	16.8 - 19.5
All Boule (CDS) Instruments	129	5.90	0.09	1.6	5.9	5.4 - 6.4	130	18.37	0.30	1.7	18.4	17.0 - 19.7
All COULTER Instruments	120	5.83	0.14	2.4	5.8	5.4 - 6.3	119	18.09	0.34	1.9	18.1	16.8 - 19.4
Abbott Cell-Dyn 1700	14	6.08	0.15	2.5	6.1	5.6 - 6.6	14	17.94	0.50	2.8	18.0	16.6 - 19.2
Abbott Cell-Dyn 1800	24	6.05	0.18	3.0	6.0	5.6 - 6.5	24	18.48	0.45	2.4	18.4	17.1 - 19.8
Abbott Cell-Dyn Emerald	71	5.77	0.12	2.0	5.8	5.3 - 6.2	72	18.06	0.27	1.5	18.1	16.7 - 19.4
Boule (CDS) Medonic M series	124	5.90	0.09	1.6	5.9	5.4 - 6.4	125	18.39	0.30	1.6	18.4	17.1 - 19.7
COULTER AcT diff/diff 2	115	5.83	0.14	2.4	5.8	5.4 - 6.3	114	18.09	0.34	1.9	18.1	16.8 - 19.4
Horiba ABX Micros/45/60	61	5.88	0.11	1.8	5.9	5.4 - 6.3	63	18.17	0.25	1.4	18.2	16.8 - 19.5

BASIC HEMATOLOGY W/ 3-PART DIFFERENTIAL–HEMOGLOBIN (g/dL) cont'd

Specimen HD-10

<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	436	5.88	0.14	2.4	5.9	5.4 - 6.3
All Abbott Cell-Dyn Instruments	108	5.89	0.18	3.1	5.8	5.4 - 6.3
All ABX Instruments	62	5.89	0.11	1.8	5.9	5.4 - 6.4
All Boule (CDS) Instruments	129	5.92	0.10	1.7	5.9	5.5 - 6.4
All COULTER Instruments	119	5.83	0.13	2.2	5.8	5.4 - 6.3
Abbott Cell-Dyn 1700	14	6.14	0.15	2.5	6.2	5.7 - 6.6
Abbott Cell-Dyn 1800	24	6.07	0.18	3.0	6.1	5.6 - 6.5
Abbott Cell-Dyn Emerald	71	5.79	0.09	1.6	5.8	5.3 - 6.2
Boule (CDS) Medonic M series	124	5.92	0.10	1.7	5.9	5.5 - 6.4
COULTER AcT diff/diff 2	114	5.83	0.13	2.1	5.8	5.4 - 6.3
Horiba ABX Micros/45/60	62	5.89	0.11	1.8	5.9	5.4 - 6.4

BASIC HEMATOLOGY W/ 3-PART DIFFERENTIAL–HEMATOCRIT (percent)

Specimen HD-6

<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	437	38.18	1.56	4.1	38.1	35.8 - 40.5
All Abbott Cell-Dyn Instruments	107	39.77	1.27	3.2	39.9	37.3 - 42.2
All ABX Instruments	63	37.02	0.85	2.3	36.9	34.7 - 39.3
All Boule (CDS) Instruments	129	36.98	0.89	2.4	37.0	34.7 - 39.2
All COULTER Instruments	122	38.62	0.97	2.5	38.6	36.2 - 41.0
Abbott Cell-Dyn 1700	12	38.89	0.85	2.2	38.9	36.5 - 41.3
Abbott Cell-Dyn 1800	24	39.85	1.19	3.0	39.8	37.4 - 42.3
Abbott Cell-Dyn Emerald	70	39.92	1.29	3.2	40.2	37.5 - 42.4
Boule (CDS) Medonic M series	124	36.98	0.89	2.4	37.0	34.7 - 39.2
COULTER AcT diff/diff 2	117	38.60	0.98	2.5	38.6	36.2 - 41.0
Horiba ABX Micros/45/60	63	37.02	0.85	2.3	36.9	34.7 - 39.3

Specimen HD-7

<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
431	50.06	1.52	3.0	50.0	47.0 - 53.1
104	51.41	1.49	2.9	51.4	48.3 - 54.5
62	49.43	1.07	2.2	49.4	46.4 - 52.4
129	49.10	1.22	2.5	49.0	46.1 - 52.1
122	50.26	1.04	2.1	50.2	47.2 - 53.3
11	50.93	1.42	2.8	51.1	47.8 - 54.0
24	51.14	1.86	3.6	51.6	48.0 - 54.3
69	51.52	1.51	2.9	51.6	48.4 - 54.7
124	49.13	1.21	2.5	49.0	46.1 - 52.1
117	50.27	1.05	2.1	50.2	47.2 - 53.3
62	49.43	1.07	2.2	49.4	46.4 - 52.4

BASIC HEMATOLOGY W/ 3-PART DIFFERENTIAL–HEMATOCRIT (percent) cont'd

<u><i>Instrument</i></u>	Specimen HD-8						Specimen HD-9					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	437	16.79	0.96	5.7	16.7	15.7 - 17.8	436	51.53	1.64	3.2	51.4	48.4 - 54.7
All Abbott Cell-Dyn Instruments	108	17.84	0.60	3.4	17.9	16.7 - 19.0	107	52.82	1.61	3.0	52.9	49.6 - 56.0
All ABX Instruments	62	15.72	0.33	2.1	15.7	14.7 - 16.7	63	49.66	0.97	2.0	49.7	46.6 - 52.7
All Boule (CDS) Instruments	129	16.06	0.36	2.2	16.0	15.0 - 17.1	130	51.47	1.25	2.4	51.5	48.3 - 54.6
All COULTER Instruments	119	17.21	0.45	2.6	17.3	16.1 - 18.3	121	51.43	1.16	2.2	51.4	48.3 - 54.6
Abbott Cell-Dyn 1700	13	17.12	0.58	3.4	17.1	16.0 - 18.2	13	51.06	1.70	3.3	51.5	47.9 - 54.2
Abbott Cell-Dyn 1800	22	18.04	0.38	2.1	18.1	16.9 - 19.2	23	52.66	1.36	2.6	52.9	49.5 - 55.9
Abbott Cell-Dyn Emerald	71	17.94	0.55	3.0	18.0	16.8 - 19.1	71	53.15	1.62	3.0	53.3	49.9 - 56.4
Boule (CDS) Medonic M series	124	16.05	0.36	2.3	16.0	15.0 - 17.1	125	51.53	1.23	2.4	51.5	48.4 - 54.7
COULTER AcT diff/diff 2	114	17.23	0.45	2.6	17.3	16.1 - 18.3	116	51.44	1.18	2.3	51.4	48.3 - 54.6
Horiba ABX Micros/45/60	62	15.72	0.33	2.1	15.7	14.7 - 16.7	63	49.66	0.97	2.0	49.7	46.6 - 52.7
Specimen HD-10												
All Method	437	16.75	0.95	5.7	16.7	15.7 - 17.8						
All Abbott Cell-Dyn Instruments	108	17.79	0.62	3.5	17.8	16.7 - 18.9						
All ABX Instruments	61	15.69	0.34	2.2	15.6	14.7 - 16.7						
All Boule (CDS) Instruments	130	15.99	0.35	2.2	16.0	15.0 - 17.0						
All COULTER Instruments	121	17.15	0.43	2.5	17.2	16.1 - 18.2						
Abbott Cell-Dyn 1700	13	17.25	0.38	2.2	17.3	16.2 - 18.3						
Abbott Cell-Dyn 1800	24	17.86	0.66	3.7	17.9	16.7 - 19.0						
Abbott Cell-Dyn Emerald	70	17.88	0.59	3.3	18.0	16.8 - 19.0						
Boule (CDS) Medonic M series	125	15.99	0.35	2.2	16.0	15.0 - 17.0						
COULTER AcT diff/diff 2	116	17.17	0.42	2.4	17.2	16.1 - 18.3						
Horiba ABX Micros/45/60	61	15.69	0.34	2.2	15.6	14.7 - 16.7						

BASIC HEMATOLOGY W/ 3-PART DIFFERENTIAL-PLATELET COUNT (x 10⁹/L)

<i><u>Instrument</u></i>	Specimen HD-6						Specimen HD-7					
	<i><u>Labs</u></i>	<i><u>Mean</u></i>	<i><u>SD</u></i>	<i><u>CV</u></i>	<i><u>Median</u></i>	<i><u>Range</u></i>	<i><u>Labs</u></i>	<i><u>Mean</u></i>	<i><u>SD</u></i>	<i><u>CV</u></i>	<i><u>Median</u></i>	<i><u>Range</u></i>
All Method	436	255.9	17.5	6.9	256	191 - 320	437	158.5	14.6	9.2	158	118 - 199
All Abbott Cell-Dyn Instruments	107	260.8	18.3	7.0	259	195 - 326	107	167.0	14.2	8.5	168	125 - 209
All ABX Instruments	61	258.5	10.9	4.2	259	193 - 324	62	156.3	10.0	6.4	156	117 - 196
All Boule (CDS) Instruments	130	240.8	13.7	5.7	240	180 - 302	130	146.5	10.4	7.1	146	109 - 184
All COULTER Instruments	122	265.6	11.4	4.3	266	199 - 333	123	164.7	12.1	7.4	164	123 - 206
Abbott Cell-Dyn 1700	14	261.2	33.0	12.6	264	195 - 327	14	170.9	17.3	10.1	173	128 - 214
Abbott Cell-Dyn 1800	24	258.0	14.6	5.6	259	193 - 323	24	171.9	14.4	8.4	171	128 - 215
Abbott Cell-Dyn Emerald	70	260.3	18.9	7.3	258	195 - 326	69	164.6	13.0	7.9	165	123 - 206
Boule (CDS) Medonic M series	125	240.2	13.4	5.6	239	180 - 301	126	145.8	9.8	6.7	146	109 - 183
COULTER AcT diff/diff 2	118	265.5	11.5	4.3	266	199 - 332	118	165.0	12.1	7.3	164	123 - 207
Horiba ABX Micros/45/60	61	258.5	10.9	4.2	259	193 - 324	62	156.3	10.0	6.4	156	117 - 196
	Specimen HD-8						Specimen HD-9					
All Method	432	67.2	7.3	10.9	66	50 - 85	438	532.6	33.2	6.2	529	399 - 666
All Abbott Cell-Dyn Instruments	106	67.8	9.2	13.6	67	50 - 85	110	539.1	42.2	7.8	537	404 - 674
All ABX Instruments	62	71.9	6.2	8.6	72	53 - 90	63	527.8	21.4	4.1	527	395 - 660
All Boule (CDS) Instruments	129	62.8	4.5	7.2	63	47 - 79	131	510.8	23.7	4.6	512	383 - 639
All COULTER Instruments	120	68.6	5.5	8.0	68	51 - 86	120	554.4	21.4	3.9	554	415 - 693
Abbott Cell-Dyn 1700	14	64.6	5.7	8.8	64	48 - 81	14	574.4	49.0	8.5	578	430 - 719
Abbott Cell-Dyn 1800	24	64.6	5.7	8.8	65	48 - 81	24	563.3	35.6	6.3	569	422 - 705
Abbott Cell-Dyn Emerald	69	70.0	11.0	15.7	69	52 - 88	72	524.1	34.9	6.7	521	393 - 656
Boule (CDS) Medonic M series	124	62.6	4.4	7.1	62	46 - 79	126	509.7	23.2	4.5	510	382 - 638
COULTER AcT diff/diff 2	115	68.8	5.4	7.9	68	51 - 86	116	553.9	21.5	3.9	553	415 - 693
Horiba ABX Micros/45/60	62	71.9	6.2	8.6	72	53 - 90	63	527.8	21.4	4.1	527	395 - 660

BASIC HEMATOLOGY W/ 3-PART DIFFERENTIAL–PLATELET COUNT (x 10⁹/L) cont'd

Specimen HD-10

<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	430	68.5	8.0	11.7	68	51 - 86
All Abbott Cell-Dyn Instruments	106	71.3	12.2	17.1	70	53 - 90
All ABX Instruments	63	73.5	5.0	6.8	74	55 - 92
All Boule (CDS) Instruments	129	63.4	5.0	7.9	63	47 - 80
All COULTER Instruments	120	69.8	5.8	8.3	70	52 - 88
Abbott Cell-Dyn 1700	14	65.6	4.6	6.9	64	49 - 82
Abbott Cell-Dyn 1800	24	66.8	6.2	9.3	67	50 - 84
Abbott Cell-Dyn Emerald	69	74.7	14.9	19.9	72	56 - 94
Boule (CDS) Medonic M series	123	63.1	4.7	7.5	63	47 - 79
COULTER AcT diff/diff 2	116	70.0	6.0	8.6	70	52 - 88
Horiba ABX Micros/45/60	63	73.5	5.0	6.8	74	55 - 92

BASIC HEMATOLOGY W/ 3-PART DIFFERENTIAL–LYMPHOCYTES (percent)

Specimen HD-6

<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	425	29.37	2.90	9.9	30.4	20.6 - 38.1
All Abbott Cell-Dyn Instruments	103	27.76	2.23	8.0	28.3	21.0 - 34.5
All ABX Instruments	60	24.57	2.68	10.9	24.4	16.5 - 32.6
All Boule (CDS) Instruments	124	30.94	0.88	2.8	30.9	28.3 - 33.6
All COULTER Instruments	115	31.22	0.90	2.9	31.3	28.5 - 34.0
Abbott Cell-Dyn 1700	15	27.98	1.37	4.9	28.5	23.8 - 32.1
Abbott Cell-Dyn 1800	23	24.27	1.33	5.5	23.9	20.2 - 28.3
Abbott Cell-Dyn Emerald	64	28.93	1.06	3.6	28.9	25.7 - 32.1
Boule (CDS) Medonic M series	124	30.94	0.88	2.8	30.9	28.3 - 33.6
COULTER AcT diff/diff 2	113	31.21	0.89	2.9	31.3	28.5 - 33.9
Horiba ABX Micros/45/60	60	24.57	2.68	10.9	24.4	16.5 - 32.6

Specimen HD-7

<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
428	56.23	6.55	11.7	57.6	36.5 - 75.9
106	52.23	3.05	5.8	53.0	43.0 - 61.4
60	45.33	4.13	9.1	44.6	32.9 - 57.8
122	63.19	1.28	2.0	63.1	59.3 - 67.1
115	58.19	1.05	1.8	58.1	55.0 - 61.4
15	53.07	2.60	4.9	53.2	45.2 - 60.9
23	47.54	1.54	3.2	47.3	42.9 - 52.2
66	53.71	1.49	2.8	53.8	49.2 - 58.2
122	63.19	1.28	2.0	63.1	59.3 - 67.1
113	58.17	1.03	1.8	58.1	55.0 - 61.3
60	45.33	4.13	9.1	44.6	32.9 - 57.8

BASIC HEMATOLOGY W/ 3-PART DIFFERENTIAL—LYMPHOCYTES (percent) cont'd

<u><i>Instrument</i></u>	Specimen HD-8						Specimen HD-9					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	424	56.99	6.73	11.8	59.8	36.7 - 77.2	422	13.54	1.19	8.8	13.7	9.9 - 17.2
All Abbott Cell-Dyn Instruments	107	53.29	3.48	6.5	53.9	42.8 - 63.8	103	13.38	1.63	12.2	13.7	8.4 - 18.3
All ABX Instruments	60	44.90	5.27	11.7	43.9	29.0 - 60.8	58	12.27	1.10	9.0	12.3	8.9 - 15.6
All Boule (CDS) Instruments	121	61.67	1.65	2.7	61.6	56.7 - 66.7	122	13.63	0.51	3.7	13.5	12.1 - 15.2
All COULTER Instruments	112	61.47	1.39	2.3	61.4	57.2 - 65.7	116	14.06	0.50	3.6	14.1	12.5 - 15.6
Abbott Cell-Dyn 1700	14	54.19	1.28	2.4	54.4	50.3 - 58.1	15	12.65	0.58	4.5	12.5	10.9 - 14.4
Abbott Cell-Dyn 1800	24	48.56	2.81	5.8	48.2	40.1 - 57.0	23	11.00	0.52	4.7	10.8	9.4 - 12.6
Abbott Cell-Dyn Emerald	68	54.66	2.26	4.1	54.7	47.8 - 61.5	64	14.40	0.95	6.6	14.2	11.5 - 17.3
Boule (CDS) Medonic M series	121	61.67	1.65	2.7	61.6	56.7 - 66.7	122	13.63	0.51	3.7	13.5	12.1 - 15.2
COULTER AcT diff/diff 2	110	61.47	1.38	2.2	61.4	57.3 - 65.6	114	14.07	0.50	3.5	14.1	12.5 - 15.6
Horiba ABX Micros/45/60	60	44.90	5.27	11.7	43.9	29.0 - 60.8	58	12.27	1.10	9.0	12.3	8.9 - 15.6
Specimen HD-10												
All Method	424	56.44	6.48	11.5	58.8	36.9 - 75.9						
All Abbott Cell-Dyn Instruments	104	52.80	3.37	6.4	53.4	42.7 - 62.9						
All ABX Instruments	60	44.99	4.98	11.1	44.2	30.0 - 60.0						
All Boule (CDS) Instruments	123	60.71	2.09	3.4	60.9	54.4 - 67.0						
All COULTER Instruments	113	61.03	1.59	2.6	61.0	56.2 - 65.9						
Abbott Cell-Dyn 1700	14	54.19	2.93	5.4	53.3	45.3 - 63.0						
Abbott Cell-Dyn 1800	24	47.85	2.59	5.4	48.0	40.0 - 55.7						
Abbott Cell-Dyn Emerald	67	54.12	2.09	3.9	54.2	47.8 - 60.4						
Boule (CDS) Medonic M series	123	60.71	2.09	3.4	60.9	54.4 - 67.0						
COULTER AcT diff/diff 2	111	60.99	1.58	2.6	61.0	56.2 - 65.8						
Horiba ABX Micros/45/60	60	44.99	4.98	11.1	44.2	30.0 - 60.0						

BASIC HEMATOLOGY W/ 3-PART DIFFERENTIAL–MONO/MID/MIXED/MCR (percent) cont'd

Specimen HD-10

<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	423	10.92	5.05	46.3	9.3	0.0 - 26.1
All Abbott Cell-Dyn Instruments	107	12.60	3.17	25.1	11.8	3.0 - 22.1
All ABX Instruments	60	20.88	3.75	18.0	20.9	9.6 - 32.2
All Boule (CDS) Instruments	121	8.59	1.82	21.2	8.9	3.1 - 14.1
All COULTER Instruments	113	7.01	1.34	19.1	7.0	2.9 - 11.1
Abbott Cell-Dyn 1700	15	12.07	1.77	14.7	12.3	6.7 - 17.4
Abbott Cell-Dyn 1800	23	17.55	1.48	8.4	17.8	13.1 - 22.0
Abbott Cell-Dyn Emerald	68	11.07	1.89	17.1	10.9	5.3 - 16.8
Boule (CDS) Medonic M series	121	8.59	1.82	21.2	8.9	3.1 - 14.1
COULTER AcT diff/diff 2	111	7.04	1.33	18.9	7.0	3.0 - 11.1
Horiba ABX Micros/45/60	60	20.88	3.75	18.0	20.9	9.6 - 32.2

BASIC HEMATOLOGY W/ 3-PART DIFFERENTIAL–GRANULOCYTES/NEUT (percent)

Specimen HD-6

<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	421	63.63	2.24	3.5	63.6	56.9 - 70.4
All Abbott Cell-Dyn Instruments	103	65.13	1.50	2.3	65.3	60.6 - 69.7
All ABX Instruments	58	66.16	1.19	1.8	66.2	62.5 - 69.8
All Boule (CDS) Instruments	122	61.46	1.40	2.3	61.4	57.2 - 65.7
All COULTER Instruments	116	63.36	1.12	1.8	63.3	59.9 - 66.8
Abbott Cell-Dyn 1700	15	63.45	1.19	1.9	63.3	59.8 - 67.1
Abbott Cell-Dyn 1800	24	64.70	1.30	2.0	65.1	60.8 - 68.6
Abbott Cell-Dyn Emerald	64	65.68	1.29	2.0	65.9	61.8 - 69.6
Boule (CDS) Medonic M series	122	61.46	1.40	2.3	61.4	57.2 - 65.7
COULTER AcT diff/diff 2	114	63.35	1.11	1.8	63.3	60.0 - 66.7
Horiba ABX Micros/45/60	58	66.16	1.19	1.8	66.2	62.5 - 69.8

Specimen HD-7

<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
425	31.31	4.86	15.5	30.8	16.7 - 45.9
106	34.86	2.90	8.3	36.2	26.1 - 43.6
60	37.66	2.32	6.2	38.3	30.6 - 44.7
122	26.04	2.14	8.2	25.9	19.6 - 32.5
116	30.27	1.54	5.1	30.3	25.6 - 34.9
15	30.61	1.88	6.2	30.5	24.9 - 36.3
24	32.35	1.33	4.1	32.4	28.3 - 36.4
65	36.87	1.16	3.1	36.8	33.4 - 40.4
122	26.04	2.14	8.2	25.9	19.6 - 32.5
115	30.29	1.54	5.1	30.3	25.6 - 34.9
60	37.66	2.32	6.2	38.3	30.6 - 44.7

HEMATOLOGY W/ 5-PART DIFFERENTIAL–WHITE BLOOD CELL COUNT (x 10⁹/L)

<i><u>Instrument</u></i>	Specimen DIF-6						Specimen DIF-7					
	<i><u>Labs</u></i>	<i><u>Mean</u></i>	<i><u>SD</u></i>	<i><u>CV</u></i>	<i><u>Median</u></i>	<i><u>Range</u></i>	<i><u>Labs</u></i>	<i><u>Mean</u></i>	<i><u>SD</u></i>	<i><u>CV</u></i>	<i><u>Median</u></i>	<i><u>Range</u></i>
All Method	14	9.57	0.12	1.3	9.6	8.1 - 11.1	14	24.32	0.57	2.3	24.3	20.6 - 28.0
All COULTER Instruments	12	9.59	0.11	1.2	9.6	8.1 - 11.1	12	24.33	0.61	2.5	24.2	20.6 - 28.0
COULTER UniCel DxH 600	10	9.54	0.11	1.2	9.5	8.1 - 11.0	10	24.30	0.41	1.7	24.3	20.6 - 28.0
Specimen DIF-8												
All Method	14	4.16	0.32	7.8	4.0	3.5 - 4.8	14	21.14	0.40	1.9	21.1	17.9 - 24.4
All COULTER Instruments	12	4.18	0.34	8.2	4.1	3.5 - 4.9	12	21.10	0.41	1.9	21.1	17.9 - 24.3
COULTER UniCel DxH 600	10	4.02	0.13	3.2	4.0	3.4 - 4.7	10	21.08	0.48	2.3	21.1	17.9 - 24.3
Specimen DIF-10												
All Method	14	4.22	0.30	7.1	4.2	3.5 - 4.9						
All COULTER Instruments	12	4.24	0.32	7.5	4.2	3.6 - 4.9						
COULTER UniCel DxH 600	10	4.14	0.09	2.2	4.2	3.5 - 4.8						

HEMATOLOGY W/ 5-PART DIFFERENTIAL–RED BLOOD CELL COUNT (x 10¹²/L)

<i><u>Instrument</u></i>	Specimen DIF-6						Specimen DIF-7					
	<i><u>Labs</u></i>	<i><u>Mean</u></i>	<i><u>SD</u></i>	<i><u>CV</u></i>	<i><u>Median</u></i>	<i><u>Range</u></i>	<i><u>Labs</u></i>	<i><u>Mean</u></i>	<i><u>SD</u></i>	<i><u>CV</u></i>	<i><u>Median</u></i>	<i><u>Range</u></i>
All Method	14	4.084	0.063	1.6	4.06	3.83 - 4.33	14	6.233	0.107	1.7	6.20	5.85 - 6.61
All COULTER Instruments	12	4.094	0.061	1.5	4.08	3.84 - 4.34	12	6.239	0.113	1.8	6.21	5.86 - 6.62
COULTER UniCel DxH 600	10	4.056	0.023	0.6	4.06	3.81 - 4.30	10	6.230	0.129	2.1	6.20	5.85 - 6.61
Specimen DIF-8												
All Method	14	2.578	0.058	2.2	2.57	2.42 - 2.74	14	5.279	0.085	1.6	5.28	4.96 - 5.60
All COULTER Instruments	12	2.576	0.062	2.4	2.57	2.42 - 2.74	12	5.279	0.091	1.7	5.28	4.96 - 5.60
COULTER UniCel DxH 600	10	2.554	0.038	1.5	2.55	2.40 - 2.71	10	5.246	0.080	1.5	5.23	4.93 - 5.57
Specimen DIF-10												
All Method	14	2.578	0.044	1.7	2.56	2.42 - 2.74						
All COULTER Instruments	12	2.580	0.047	1.8	2.57	2.42 - 2.74						
COULTER UniCel DxH 600	10	2.572	0.037	1.4	2.58	2.41 - 2.73						

HEMATOLOGY W/ 5-PART DIFFERENTIAL–HEMOGLOBIN (g/dL)

<u>Instrument</u>	Specimen DIF-6						Specimen DIF-7					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	14	11.76	0.18	1.5	11.8	10.9 - 12.6	14	19.69	0.17	0.9	19.7	18.3 - 21.1
All COULTER Instruments	12	11.78	0.18	1.6	11.8	10.9 - 12.6	12	19.71	0.16	0.8	19.7	18.3 - 21.1
COULTER UniCel DxH 600	10	11.74	0.23	2.0	11.7	10.9 - 12.6	10	19.70	0.14	0.7	19.7	18.3 - 21.1
Specimen DIF-8						Specimen DIF-9						
All Method	14	5.90	0.09	1.5	5.9	5.4 - 6.4	14	16.98	0.22	1.3	17.0	15.7 - 18.2
All COULTER Instruments	12	5.91	0.08	1.4	5.9	5.4 - 6.4	12	17.03	0.18	1.1	17.1	15.8 - 18.3
COULTER UniCel DxH 600	10	5.90	0.07	1.2	5.9	5.4 - 6.4	10	16.98	0.13	0.8	17.0	15.7 - 18.2
Specimen DIF-10												
All Method	14	5.91	0.12	2.0	5.9	5.4 - 6.4						
All COULTER Instruments	12	5.93	0.12	2.0	6.0	5.5 - 6.4						
COULTER UniCel DxH 600	10	5.90	0.10	1.7	5.9	5.4 - 6.4						

HEMATOLOGY W/ 5-PART DIFFERENTIAL–HEMATOCRIT (percent)

<u>Instrument</u>	Specimen DIF-6						Specimen DIF-7					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	14	36.21	0.31	0.9	36.3	34.0 - 38.4	14	63.38	1.17	1.8	63.2	59.5 - 67.2
All COULTER Instruments	12	36.28	0.26	0.7	36.4	34.0 - 38.5	12	63.39	1.25	2.0	63.2	59.5 - 67.2
COULTER UniCel DxH 600	10	36.32	0.13	0.4	36.4	34.1 - 38.5	10	63.90	1.21	1.9	63.5	60.0 - 67.8
Specimen DIF-8						Specimen DIF-9						
All Method	14	18.87	0.35	1.9	19.0	17.7 - 20.0	14	53.63	0.81	1.5	53.8	50.4 - 56.9
All COULTER Instruments	12	18.85	0.37	2.0	18.9	17.7 - 20.0	12	53.61	0.86	1.6	53.6	50.3 - 56.9
COULTER UniCel DxH 600	10	18.90	0.24	1.3	18.8	17.7 - 20.1	10	53.72	0.87	1.6	53.3	50.4 - 57.0
Specimen DIF-10												
All Method	14	18.86	0.34	1.8	18.9	17.7 - 20.0						
All COULTER Instruments	12	18.86	0.37	1.9	18.9	17.7 - 20.0						
COULTER UniCel DxH 600	10	19.02	0.24	1.3	19.0	17.8 - 20.2						

HEMATOLOGY W/ 5-PART DIFFERENTIAL-PLATELET COUNT (x 10⁹/L)

<u><i>Instrument</i></u>	Specimen DIF-6						Specimen DIF-7					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	14	253.8	5.7	2.2	253	190 - 318	14	464.3	15.1	3.3	464	348 - 581
All COULTER Instruments	12	253.9	6.0	2.4	255	190 - 318	12	464.4	16.2	3.5	464	348 - 581
COULTER UniCel DxH 600	10	252.6	7.4	2.9	251	189 - 316	10	467.2	9.7	2.1	468	350 - 584
<u><i>Instrument</i></u>	Specimen DIF-8						Specimen DIF-9					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	14	80.8	3.1	3.8	81	60 - 101	14	486.3	9.2	1.9	488	364 - 608
All COULTER Instruments	12	80.8	3.3	4.1	80	60 - 101	12	488.5	7.0	1.4	489	366 - 611
COULTER UniCel DxH 600	10	80.0	2.5	3.2	79	60 - 100	10	489.4	2.8	0.6	488	367 - 612
<u><i>Instrument</i></u>	Specimen DIF-10											
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>						
All Method	14	80.6	4.8	6.0	79	60 - 101						
All COULTER Instruments	12	80.9	5.1	6.3	80	60 - 102						
COULTER UniCel DxH 600	10	79.2	2.9	3.6	79	59 - 99						

HEMATOLOGY W/ 5-PART DIFFERENTIAL-NEUTROPHILS (percent)

<u><i>Instrument</i></u>	Specimen DIF-6						Specimen DIF-7					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	14	57.49	1.28	2.2	57.8	53.6 - 61.4	14	68.43	0.45	0.7	68.3	67.0 - 69.8
All COULTER Instruments	12	57.35	1.29	2.3	57.6	53.4 - 61.3	12	68.36	0.42	0.6	68.3	67.0 - 69.7
COULTER UniCel DxH 600	10	57.08	1.08	1.9	57.4	53.8 - 60.4	10	68.52	0.36	0.5	68.3	67.4 - 69.6
<u><i>Instrument</i></u>	Specimen DIF-8						Specimen DIF-9					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	14	50.34	1.22	2.4	50.4	46.6 - 54.0	14	67.67	0.81	1.2	67.6	65.2 - 70.1
All COULTER Instruments	12	50.14	1.12	2.2	50.3	46.7 - 53.5	12	67.60	0.84	1.2	67.6	65.0 - 70.2
COULTER UniCel DxH 600	10	50.08	1.23	2.5	50.1	46.3 - 53.8	10	67.38	0.90	1.3	67.5	64.6 - 70.1
<u><i>Instrument</i></u>	Specimen DIF-10											
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>						
All Method	14	50.79	0.48	0.9	50.7	49.3 - 52.3						
All COULTER Instruments	12	50.81	0.50	1.0	50.8	49.3 - 52.4						
COULTER UniCel DxH 600	10	50.86	0.53	1.0	50.9	49.2 - 52.5						

HEMATOLOGY W/ 5-PART DIFFERENTIAL– LYMPHOCYTES (percent)

<i><u>Instrument</u></i>	Specimen DIF-6						Specimen DIF-7					
	<i><u>Labs</u></i>	<i><u>Mean</u></i>	<i><u>SD</u></i>	<i><u>CV</u></i>	<i><u>Median</u></i>	<i><u>Range</u></i>	<i><u>Labs</u></i>	<i><u>Mean</u></i>	<i><u>SD</u></i>	<i><u>CV</u></i>	<i><u>Median</u></i>	<i><u>Range</u></i>
All Method	14	30.82	0.75	2.4	31.0	28.5 - 33.1	14	21.60	0.48	2.2	21.5	20.1 - 23.1
All COULTER Instruments	12	30.98	0.64	2.1	31.1	29.0 - 32.9	12	21.55	0.49	2.3	21.5	20.0 - 23.1
COULTER UniCel DxH 600	10	30.98	0.74	2.4	31.0	28.7 - 33.2	10	21.44	0.38	1.8	21.4	20.2 - 22.6
<i><u>Instrument</u></i>	Specimen DIF-8						Specimen DIF-9					
	<i><u>Labs</u></i>	<i><u>Mean</u></i>	<i><u>SD</u></i>	<i><u>CV</u></i>	<i><u>Median</u></i>	<i><u>Range</u></i>	<i><u>Labs</u></i>	<i><u>Mean</u></i>	<i><u>SD</u></i>	<i><u>CV</u></i>	<i><u>Median</u></i>	<i><u>Range</u></i>
All Method	14	40.49	1.23	3.0	40.3	36.8 - 44.2	14	21.92	0.90	4.1	21.9	19.2 - 24.7
All COULTER Instruments	12	40.74	1.04	2.6	40.7	37.6 - 43.9	12	22.00	0.93	4.2	21.9	19.2 - 24.8
COULTER UniCel DxH 600	10	40.88	1.30	3.2	41.1	36.9 - 44.8	10	22.32	0.95	4.3	21.9	19.4 - 25.2
<i><u>Instrument</u></i>	Specimen DIF-10											
	<i><u>Labs</u></i>	<i><u>Mean</u></i>	<i><u>SD</u></i>	<i><u>CV</u></i>	<i><u>Median</u></i>	<i><u>Range</u></i>						
All Method	14	40.09	0.44	1.1	40.1	38.7 - 41.5						
All COULTER Instruments	12	40.06	0.46	1.1	40.1	38.6 - 41.5						
COULTER UniCel DxH 600	10	40.08	0.49	1.2	40.1	38.6 - 41.6						

HEMATOLOGY W/ 5-PART DIFFERENTIAL– MONOCYTES (percent)

<i><u>Instrument</u></i>	Specimen DIF-6						Specimen DIF-7					
	<i><u>Labs</u></i>	<i><u>Mean</u></i>	<i><u>SD</u></i>	<i><u>CV</u></i>	<i><u>Median</u></i>	<i><u>Range</u></i>	<i><u>Labs</u></i>	<i><u>Mean</u></i>	<i><u>SD</u></i>	<i><u>CV</u></i>	<i><u>Median</u></i>	<i><u>Range</u></i>
All Method	14	5.23	0.78	14.9	5.3	2.8 - 7.6	14	5.83	0.41	7.0	5.9	4.6 - 7.1
All COULTER Instruments	12	5.23	0.83	15.9	5.3	2.7 - 7.8	12	5.89	0.40	6.8	6.0	4.6 - 7.1
COULTER UniCel DxH 600	10	5.56	0.58	10.4	5.3	3.8 - 7.3	10	5.98	0.18	3.0	6.0	5.4 - 6.6
<i><u>Instrument</u></i>	Specimen DIF-8						Specimen DIF-9					
	<i><u>Labs</u></i>	<i><u>Mean</u></i>	<i><u>SD</u></i>	<i><u>CV</u></i>	<i><u>Median</u></i>	<i><u>Range</u></i>	<i><u>Labs</u></i>	<i><u>Mean</u></i>	<i><u>SD</u></i>	<i><u>CV</u></i>	<i><u>Median</u></i>	<i><u>Range</u></i>
All Method	14	1.22	0.39	31.6	1.4	0.0 - 2.4	14	6.08	0.56	9.2	6.1	4.4 - 7.8
All COULTER Instruments	12	1.18	0.38	32.7	1.3	0.0 - 2.4	12	6.03	0.57	9.5	6.1	4.3 - 7.8
COULTER UniCel DxH 600	10	1.32	0.26	19.6	1.4	0.5 - 2.1	10	6.20	0.19	3.0	6.1	5.6 - 6.8
<i><u>Instrument</u></i>	Specimen DIF-10											
	<i><u>Labs</u></i>	<i><u>Mean</u></i>	<i><u>SD</u></i>	<i><u>CV</u></i>	<i><u>Median</u></i>	<i><u>Range</u></i>						
All Method	14	1.24	0.32	25.8	1.3	0.2 - 2.3						
All COULTER Instruments	12	1.21	0.33	27.0	1.2	0.2 - 2.2						
COULTER UniCel DxH 600	10	1.28	0.18	14.0	1.3	0.7 - 1.9						

HEMATOLOGY W/ 5-PART DIFFERENTIAL– EOSINOPHILS (percent)

<u><i>Instrument</i></u>	Specimen DIF-6						Specimen DIF-7					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	13	6.25	0.36	5.8	6.2	5.1 - 7.4	14	4.06	0.22	5.4	4.1	3.3 - 4.8
All COULTER Instruments	11	6.23	0.39	6.2	6.1	5.0 - 7.4	12	4.09	0.21	5.1	4.1	3.4 - 4.8
COULTER UniCel DxH 600	9	-	-	-	6.2	5.0 - 7.4	10	4.02	0.16	4.1	4.1	3.5 - 4.6
Specimen DIF-8						Specimen DIF-9						
All Method	14	7.89	0.46	5.9	7.7	6.4 - 9.3	14	4.27	0.30	7.0	4.2	3.3 - 5.2
All COULTER Instruments	12	7.90	0.50	6.3	7.7	6.4 - 9.4	12	4.30	0.30	7.0	4.3	3.3 - 5.3
COULTER UniCel DxH 600	10	7.68	0.29	3.7	7.7	6.8 - 8.6	10	4.16	0.27	6.5	4.1	3.3 - 5.0
Specimen DIF-10												
All Method	14	7.83	0.42	5.4	7.8	6.5 - 9.1						
All COULTER Instruments	12	7.86	0.44	5.6	7.9	6.5 - 9.2						
COULTER UniCel DxH 600	10	7.74	0.49	6.4	7.6	6.2 - 9.3						

HEMATOLOGY W/ 5-PART DIFFERENTIAL– BASOPHILS (percent)

<u><i>Instrument</i></u>	Specimen DIF-6						Specimen DIF-7					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	14	0.16	0.20	129.0	0.1	0.0 - 0.8	14	0.10	0.15	150.0	0.0	0.0 - 0.6
All COULTER Instruments	12	0.16	0.21	131.3	0.1	0.0 - 0.9	12	0.11	0.16	138.0	0.1	0.0 - 0.6
COULTER UniCel DxH 600	10	0.04	0.05	136.9	0.0	0.0 - 0.3	10	0.04	0.05	136.9	0.0	0.0 - 0.3
Specimen DIF-8						Specimen DIF-9						
All Method	14	0.06	0.05	94.9	0.1	0.0 - 0.3	14	0.06	0.05	94.9	0.1	0.0 - 0.3
All COULTER Instruments	12	0.05	0.05	106.9	0.1	0.0 - 0.3	12	0.05	0.05	106.9	0.1	0.0 - 0.3
COULTER UniCel DxH 600	10	0.04	0.05	136.9	0.0	0.0 - 0.3	10	0.04	0.05	136.9	0.0	0.0 - 0.3
Specimen DIF-10												
All Method	14	0.04	0.05	118.6	0.0	0.0 - 0.3						
All COULTER Instruments	12	0.05	0.05	106.9	0.1	0.0 - 0.3						
COULTER UniCel DxH 600	10	0.04	0.05	136.9	0.0	0.0 - 0.3						

BLOOD LEAD (µg/dL)

<u>Instrument</u>	Specimen LED-6						Specimen LED-7					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	19	5.05	0.77	15.3	4.9	1.0 - 9.1	20	25.97	1.87	7.2	26.0	21.9 - 30.0
All Magellan Diagnostics Methods	19	5.05	0.77	15.3	4.9	1.0 - 9.1	20	25.97	1.87	7.2	26.0	21.9 - 30.0
Magellan Diagnostics LeadCare II	19	5.05	0.77	15.3	4.9	1.0 - 9.1	20	25.97	1.87	7.2	26.0	21.9 - 30.0

<u>Instrument</u>	Specimen LED-8						Specimen LED-9					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	5	45.15	1.34	3.0	45.2	40.6 - 49.7	5	19.75	2.47	12.5	19.8	15.7 - 23.8
All Magellan Diagnostics Methods	5	45.15	1.34	3.0	45.2	40.6 - 49.7	5	19.75	2.47	12.5	19.8	15.7 - 23.8

<u>Instrument</u>	Specimen LED-10					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	5	45.00	1.13	2.5	45.0	40.5 - 49.5
All Magellan Diagnostics Methods	5	45.00	1.13	2.5	45.0	40.5 - 49.5

RETICULOCYTE COUNT (percent)

<u>Instrument</u>	Specimen RT-3						Specimen RT-4					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	24	3.75	0.41	10.9	3.8	2.6 - 4.9	24	1.20	0.20	16.5	1.3	0.8 - 1.7
All Automated Methods	14	3.64	0.34	9.3	3.7	2.5 - 4.8	14	1.29	0.15	11.4	1.3	0.9 - 1.7
All Manual Methods	10	4.31	1.31	30.4	3.9	1.6 - 7.0	10	1.35	0.95	70.3	1.1	0.0 - 3.3

HEMATOLOGY W/ 5-PART DIFFERENTIAL-PLATELET COUNT (x 10⁹/L)

<u><i>Instrument</i></u>	Specimen BCX-6						Specimen BCX-7					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	96	475.4	16.9	3.5	475	356 - 595	96	74.1	4.6	6.2	74	55 - 93
All ABX Instruments	88	476.4	16.3	3.4	476	357 - 596	88	74.0	4.6	6.2	74	55 - 93
All COULTER Instruments	10	466.8	21.1	4.5	470	350 - 584	10	75.8	5.0	6.6	75	56 - 95
ABX Pentra 60C+	78	476.4	16.7	3.5	476	357 - 596	78	73.8	4.7	6.3	74	55 - 93
ABX Pentra 80 / XL 80	10	478.0	14.1	3.0	485	358 - 598	10	76.9	2.0	2.6	76	57 - 97
COULTER AcT 5diff	10	466.8	21.1	4.5	470	350 - 584	10	75.8	5.0	6.6	75	56 - 95
	Specimen BCX-8						Specimen BCX-9					
All Method	96	285.8	11.5	4.0	286	214 - 358	96	490.4	13.9	2.8	490	367 - 613
All ABX Instruments	88	286.3	11.5	4.0	286	214 - 358	88	491.6	13.8	2.8	491	368 - 615
All COULTER Instruments	10	281.6	11.1	3.9	283	211 - 352	10	479.7	11.0	2.3	482	359 - 600
ABX Pentra 60C+	78	286.5	11.8	4.1	286	214 - 359	78	492.0	13.6	2.8	491	369 - 616
ABX Pentra 80 / XL 80	10	284.0	10.3	3.6	288	213 - 355	10	484.7	14.2	2.9	485	363 - 606
COULTER AcT 5diff	10	281.6	11.1	3.9	283	211 - 352	10	479.7	11.0	2.3	482	359 - 600
	Specimen BCX-10											
All Method	96	73.2	4.9	6.6	73	54 - 92						
All ABX Instruments	88	73.4	5.0	6.8	73	55 - 92						
All COULTER Instruments	10	71.9	2.5	3.5	73	53 - 90						
ABX Pentra 60C+	78	73.8	4.9	6.6	74	55 - 93						
ABX Pentra 80 / XL 80	10	68.3	4.3	6.4	68	51 - 86						
COULTER AcT 5diff	10	71.9	2.5	3.5	73	53 - 90						

HEMATOLOGY W/ 5 or 6-PART DIFFERENTIAL-RED BLOOD CELL COUNT (x 10¹²/L)

<u>Instrument</u>	Specimen MX-6						Specimen MX-7					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	110	4.668	0.044	0.9	4.67	4.38 - 4.95	110	6.248	0.065	1.0	6.25	5.87 - 6.63
All Sysmex XE/XT Instruments	10	4.702	0.059	1.3	4.68	4.41 - 4.99	10	6.174	0.123	2.0	6.13	5.80 - 6.55
All Sysmex XN/XS Instruments	105	4.667	0.043	0.9	4.66	4.38 - 4.95	105	6.249	0.064	1.0	6.25	5.87 - 6.63
Sysmex XN-1000	14	4.672	0.026	0.6	4.67	4.39 - 4.96	14	6.240	0.059	0.9	6.24	5.86 - 6.62
Sysmex XN-430	27	4.675	0.066	1.4	4.68	4.39 - 4.96	27	6.273	0.065	1.0	6.28	5.89 - 6.65
Sysmex XN-550	17	4.660	0.043	0.9	4.66	4.38 - 4.94	17	6.244	0.043	0.7	6.24	5.86 - 6.62
Sysmex XS-1000i	36	4.664	0.044	0.9	4.67	4.38 - 4.95	36	6.227	0.068	1.1	6.23	5.85 - 6.61
<u>Instrument</u>	Specimen MX-8						Specimen MX-9					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	110	2.195	0.033	1.5	2.20	2.06 - 2.33	110	5.791	0.065	1.1	5.79	5.44 - 6.14
All Sysmex XE/XT Instruments	10	2.298	0.032	1.4	2.32	2.16 - 2.44	10	5.692	0.170	3.0	5.70	5.35 - 6.04
All Sysmex XN/XS Instruments	105	2.194	0.031	1.4	2.20	2.06 - 2.33	105	5.793	0.063	1.1	5.79	5.44 - 6.15
Sysmex XN-1000	14	2.195	0.023	1.0	2.20	2.06 - 2.33	14	5.794	0.050	0.9	5.78	5.44 - 6.15
Sysmex XN-430	27	2.187	0.032	1.5	2.19	2.05 - 2.32	27	5.816	0.066	1.1	5.81	5.46 - 6.17
Sysmex XN-550	17	2.178	0.023	1.1	2.19	2.04 - 2.31	17	5.803	0.060	1.0	5.80	5.45 - 6.16
Sysmex XS-1000i	36	2.214	0.024	1.1	2.21	2.08 - 2.35	36	5.768	0.060	1.0	5.77	5.42 - 6.12
<u>Instrument</u>	Specimen MX-10											
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>						
All Method	110	2.196	0.035	1.6	2.20	2.06 - 2.33						
All Sysmex XE/XT Instruments	10	2.292	0.026	1.1	2.30	2.15 - 2.43						
All Sysmex XN/XS Instruments	105	2.194	0.030	1.3	2.20	2.06 - 2.33						
Sysmex XN-1000	14	2.186	0.026	1.2	2.18	2.05 - 2.32						
Sysmex XN-430	27	2.183	0.030	1.4	2.18	2.05 - 2.32						
Sysmex XN-550	17	2.177	0.030	1.4	2.17	2.04 - 2.31						
Sysmex XS-1000i	36	2.214	0.025	1.1	2.22	2.08 - 2.35						

HEMATOLOGY W/ 5 or 6-PART DIFFERENTIAL–HEMOGLOBIN (g/dL)

<u><i>Instrument</i></u>	Specimen MX-6						Specimen MX-7					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	110	13.22	0.12	0.9	13.2	12.2 - 14.2	110	19.11	0.22	1.1	19.1	17.7 - 20.5
All Sysmex XE/XT Instruments	10	13.28	0.13	1.0	13.2	12.3 - 14.3	10	19.02	0.26	1.4	19.1	17.6 - 20.4
All Sysmex XN/XS Instruments	105	13.22	0.12	0.9	13.2	12.2 - 14.2	105	19.12	0.22	1.1	19.1	17.7 - 20.5
Sysmex XN-1000	14	13.23	0.13	1.0	13.2	12.3 - 14.2	14	19.04	0.20	1.1	19.1	17.7 - 20.4
Sysmex XN-430	27	13.17	0.13	1.0	13.2	12.2 - 14.1	27	19.00	0.18	0.9	19.0	17.6 - 20.4
Sysmex XN-550	17	13.21	0.09	0.7	13.2	12.2 - 14.2	17	19.07	0.15	0.8	19.0	17.7 - 20.5
Sysmex XS-1000i	36	13.28	0.11	0.9	13.3	12.3 - 14.3	36	19.28	0.19	1.0	19.3	17.9 - 20.7

<u><i>Instrument</i></u>	Specimen MX-8						Specimen MX-9					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	110	5.75	0.07	1.2	5.8	5.3 - 6.2	110	17.71	0.18	1.0	17.7	16.4 - 19.0
All Sysmex XE/XT Instruments	10	5.86	0.09	1.5	5.8	5.4 - 6.3	10	17.74	0.24	1.4	17.8	16.4 - 19.0
All Sysmex XN/XS Instruments	105	5.75	0.07	1.2	5.8	5.3 - 6.2	105	17.71	0.18	1.0	17.7	16.4 - 19.0
Sysmex XN-1000	14	5.76	0.08	1.3	5.8	5.3 - 6.2	14	17.69	0.15	0.8	17.7	16.4 - 19.0
Sysmex XN-430	27	5.76	0.06	1.1	5.8	5.3 - 6.2	27	17.59	0.12	0.7	17.6	16.3 - 18.9
Sysmex XN-550	17	5.78	0.07	1.3	5.8	5.3 - 6.2	17	17.66	0.13	0.8	17.6	16.4 - 18.9
Sysmex XS-1000i	36	5.73	0.07	1.3	5.7	5.3 - 6.2	36	17.84	0.17	1.0	17.8	16.5 - 19.1

<u><i>Instrument</i></u>	Specimen MX-10					
	<u><i>Labs</i></u>	<u><i>Mean</i></u>	<u><i>SD</i></u>	<u><i>CV</i></u>	<u><i>Median</i></u>	<u><i>Range</i></u>
All Method	110	5.75	0.08	1.3	5.8	5.3 - 6.2
All Sysmex XE/XT Instruments	10	5.86	0.09	1.5	5.8	5.4 - 6.3
All Sysmex XN/XS Instruments	105	5.75	0.07	1.3	5.7	5.3 - 6.2
Sysmex XN-1000	14	5.74	0.10	1.8	5.8	5.3 - 6.2
Sysmex XN-430	27	5.75	0.07	1.2	5.8	5.3 - 6.2
Sysmex XN-550	17	5.77	0.07	1.2	5.8	5.3 - 6.2
Sysmex XS-1000i	36	5.74	0.07	1.2	5.7	5.3 - 6.2

HEMATOLOGY W/ 5 or 6-PART DIFFERENTIAL – HEMATOCRIT (percent)

<i>Instrument</i>	Specimen MX-6						Specimen MX-7					
	<i>Labs</i>	<i>Mean</i>	<i>SD</i>	<i>CV</i>	<i>Median</i>	<i>Range</i>	<i>Labs</i>	<i>Mean</i>	<i>SD</i>	<i>CV</i>	<i>Median</i>	<i>Range</i>
All Method	110	39.50	0.66	1.7	39.5	37.1 - 41.9	110	56.93	1.02	1.8	57.1	53.5 - 60.4
All Sysmex XE/XT Instruments	10	39.90	0.69	1.7	40.0	37.5 - 42.3	10	56.90	1.03	1.8	56.9	53.4 - 60.4
All Sysmex XN/XS Instruments	105	39.48	0.65	1.6	39.5	37.1 - 41.9	105	56.93	1.02	1.8	57.1	53.5 - 60.4
Sysmex XN-1000	14	39.43	0.47	1.2	39.4	37.0 - 41.8	14	56.83	0.62	1.1	57.0	53.4 - 60.3
Sysmex XN-430	27	39.25	0.86	2.2	39.4	36.8 - 41.7	27	56.68	1.08	1.9	56.7	53.2 - 60.1
Sysmex XN-550	17	39.25	0.74	1.9	39.2	36.8 - 41.7	17	56.62	0.96	1.7	56.5	53.2 - 60.1
Sysmex XS-1000i	36	39.73	0.55	1.4	39.9	37.3 - 42.2	36	57.25	1.09	1.9	57.3	53.8 - 60.7
<i>Instrument</i>	Specimen MX-8						Specimen MX-9					
	<i>Labs</i>	<i>Mean</i>	<i>SD</i>	<i>CV</i>	<i>Median</i>	<i>Range</i>	<i>Labs</i>	<i>Mean</i>	<i>SD</i>	<i>CV</i>	<i>Median</i>	<i>Range</i>
All Method	110	17.45	0.43	2.5	17.5	16.4 - 18.6	110	52.91	0.99	1.9	53.0	49.7 - 56.1
All Sysmex XE/XT Instruments	10	18.22	0.40	2.2	18.4	17.1 - 19.4	10	52.52	1.91	3.6	52.6	49.3 - 55.7
All Sysmex XN/XS Instruments	105	17.43	0.39	2.2	17.5	16.3 - 18.5	105	52.90	0.99	1.9	53.0	49.7 - 56.1
Sysmex XN-1000	14	17.26	0.23	1.4	17.3	16.2 - 18.3	14	52.76	0.54	1.0	52.8	49.5 - 56.0
Sysmex XN-430	27	17.24	0.38	2.2	17.3	16.2 - 18.3	27	52.62	1.02	1.9	52.7	49.4 - 55.8
Sysmex XN-550	17	17.26	0.35	2.0	17.3	16.2 - 18.3	17	52.71	1.20	2.3	52.6	49.5 - 55.9
Sysmex XS-1000i	36	17.76	0.26	1.5	17.8	16.6 - 18.9	36	53.27	0.94	1.8	53.4	50.0 - 56.5
<i>Instrument</i>	Specimen MX-10											
	<i>Labs</i>	<i>Mean</i>	<i>SD</i>	<i>CV</i>	<i>Median</i>	<i>Range</i>						
All Method	110	17.42	0.43	2.5	17.5	16.3 - 18.5						
All Sysmex XE/XT Instruments	10	18.22	0.30	1.7	18.4	17.1 - 19.4						
All Sysmex XN/XS Instruments	105	17.39	0.40	2.3	17.4	16.3 - 18.5						
Sysmex XN-1000	14	17.15	0.24	1.4	17.2	16.1 - 18.2						
Sysmex XN-430	27	17.20	0.40	2.3	17.3	16.1 - 18.3						
Sysmex XN-550	17	17.21	0.36	2.1	17.1	16.1 - 18.3						
Sysmex XS-1000i	36	17.73	0.22	1.2	17.7	16.6 - 18.8						

HEMATOLOGY W/ 5 or 6-PART DIFFERENTIAL-PLATELET COUNT (x 10⁹/L)

Specimen MX-6							Specimen MX-7					
<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	110	217.3	7.9	3.6	217	162 - 272	110	415.2	19.3	4.6	419	311 - 520
All Sysmex XE/XT Instruments	10	210.6	10.0	4.7	210	157 - 264	10	388.8	20.3	5.2	381	291 - 486
All Sysmex XN/XS Instruments	105	217.6	7.7	3.5	218	163 - 273	105	416.5	18.4	4.4	420	312 - 521
Sysmex XN-1000	14	213.4	6.6	3.1	215	160 - 267	14	417.7	16.3	3.9	422	313 - 523
Sysmex XN-430	27	222.5	7.8	3.5	222	166 - 279	27	427.0	12.7	3.0	425	320 - 534
Sysmex XN-550	17	221.2	8.0	3.6	220	165 - 277	17	426.4	10.8	2.5	425	319 - 534
Sysmex XS-1000i	36	215.1	6.2	2.9	215	161 - 269	36	398.4	10.5	2.6	396	298 - 499
Specimen MX-8							Specimen MX-9					
All Method	110	55.3	3.7	6.6	55	41 - 70	110	413.4	18.1	4.4	413	310 - 517
All Sysmex XE/XT Instruments	10	56.6	9.0	15.8	56	42 - 71	10	383.2	25.6	6.7	384	287 - 479
All Sysmex XN/XS Instruments	105	55.3	3.5	6.3	55	41 - 70	105	414.3	17.4	4.2	414	310 - 518
Sysmex XN-1000	14	52.6	2.6	5.0	52	39 - 66	14	411.8	13.4	3.2	410	308 - 515
Sysmex XN-430	27	54.5	3.7	6.7	54	40 - 69	27	420.8	13.4	3.2	417	315 - 527
Sysmex XN-550	17	55.4	3.1	5.6	55	41 - 70	17	427.4	11.7	2.7	428	320 - 535
Sysmex XS-1000i	36	56.7	3.0	5.4	57	42 - 71	36	400.8	11.9	3.0	401	300 - 501
Specimen MX-10												
All Method	110	55.3	3.9	7.1	55	41 - 70						
All Sysmex XE/XT Instruments	10	58.6	7.6	13.0	59	43 - 74						
All Sysmex XN/XS Instruments	105	55.2	3.7	6.7	55	41 - 69						
Sysmex XN-1000	14	52.5	2.7	5.1	52	39 - 66						
Sysmex XN-430	27	54.3	3.7	6.9	54	40 - 68						
Sysmex XN-550	17	55.2	3.1	5.6	55	41 - 69						
Sysmex XS-1000i	36	57.6	3.8	6.6	57	43 - 72						

HEMATOLOGY W/ 5 or 6-PART DIFFERENTIAL – NEUTROPHILS (percent)

Specimen MX-6							Specimen MX-7					
<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	110	49.91	2.02	4.1	49.6	43.8 - 56.0	110	58.49	1.71	2.9	58.0	53.3 - 63.7
All Sysmex XE/XT Instruments	10	52.48	0.38	0.7	52.4	51.3 - 53.7	10	59.94	1.82	3.0	60.0	54.4 - 65.5
All Sysmex XN/XS Instruments	105	49.77	1.98	4.0	49.5	43.8 - 55.8	105	58.42	1.67	2.9	58.0	53.3 - 63.5
Sysmex XN-1000	14	51.79	1.06	2.0	52.3	48.6 - 55.0	14	60.69	1.33	2.2	61.1	56.6 - 64.7
Sysmex XN-430	27	48.77	1.10	2.3	48.8	45.4 - 52.1	27	58.05	1.08	1.9	57.7	54.8 - 61.3
Sysmex XN-550	17	51.13	4.79	9.4	49.5	36.7 - 65.5	17	57.46	0.85	1.5	57.7	54.9 - 60.1
Sysmex XS-1000i	36	49.87	2.72	5.5	49.5	41.6 - 58.1	36	57.94	1.09	1.9	58.1	54.6 - 61.3
Specimen MX-8							Specimen MX-9					
All Method	110	61.08	2.02	3.3	60.9	55.0 - 67.2	110	58.32	1.73	3.0	58.0	53.1 - 63.6
All Sysmex XE/XT Instruments	10	64.32	1.67	2.6	64.5	59.3 - 69.4	10	59.38	2.29	3.9	59.0	52.5 - 66.3
All Sysmex XN/XS Instruments	105	60.91	1.90	3.1	60.8	55.2 - 66.7	105	58.26	1.70	2.9	58.0	53.1 - 63.4
Sysmex XN-1000	14	62.71	0.85	1.4	62.9	60.1 - 65.3	14	60.76	0.73	1.2	61.0	58.5 - 63.0
Sysmex XN-430	27	60.23	1.62	2.7	60.0	55.3 - 65.1	27	58.00	1.75	3.0	57.7	52.7 - 63.3
Sysmex XN-550	17	60.31	1.46	2.4	59.9	55.9 - 64.8	17	57.74	1.69	2.9	58.0	52.6 - 62.9
Sysmex XS-1000i	36	61.15	2.00	3.3	60.9	55.1 - 67.2	36	57.79	1.17	2.0	57.8	54.2 - 61.4
Specimen MX-10												
All Method	110	60.89	1.97	3.2	60.5	54.9 - 66.9						
All Sysmex XE/XT Instruments	10	63.92	1.79	2.8	63.6	58.5 - 69.3						
All Sysmex XN/XS Instruments	105	60.73	1.86	3.1	60.5	55.1 - 66.4						
Sysmex XN-1000	14	62.59	1.79	2.9	63.0	57.2 - 68.0						
Sysmex XN-430	27	60.33	1.95	3.2	59.9	54.4 - 66.2						
Sysmex XN-550	17	60.16	1.71	2.8	59.9	55.0 - 65.3						
Sysmex XS-1000i	36	60.66	1.46	2.4	60.5	56.2 - 65.1						

HEMATOLOGY W/ 5 or 6-PART DIFFERENTIAL – LYMPHOCYTES (percent)

<u>Instrument</u>	Specimen MX-6						Specimen MX-7					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	110	28.45	2.29	8.0	28.6	21.5 - 35.4	110	17.41	1.05	6.1	17.5	14.2 - 20.6
All Sysmex XE/XT Instruments	10	31.20	3.69	11.8	32.6	20.1 - 42.3	10	22.60	3.55	15.7	22.6	11.9 - 33.3
All Sysmex XN/XS Instruments	105	28.30	2.12	7.5	28.6	21.9 - 34.7	105	17.49	0.75	4.3	17.5	15.2 - 19.8
Sysmex XN-1000	14	26.64	0.88	3.3	26.5	23.9 - 29.3	14	17.58	1.51	8.6	17.1	13.0 - 22.2
Sysmex XN-430	27	29.31	0.97	3.3	29.4	26.3 - 32.3	27	17.52	0.64	3.6	17.6	15.6 - 19.5
Sysmex XN-550	17	29.09	0.81	2.8	29.0	26.6 - 31.6	17	17.50	0.40	2.3	17.6	16.3 - 18.7
Sysmex XS-1000i	36	27.98	3.00	10.7	28.2	18.9 - 37.0	36	17.56	0.56	3.2	17.5	15.8 - 19.3
<u>Instrument</u>	Specimen MX-8						Specimen MX-9					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	110	14.21	1.97	13.9	14.5	8.2 - 20.2	110	17.58	1.14	6.5	17.7	14.1 - 21.0
All Sysmex XE/XT Instruments	10	17.56	2.37	13.5	18.6	10.4 - 24.7	10	22.64	3.10	13.7	23.3	13.3 - 32.0
All Sysmex XN/XS Instruments	105	14.15	1.62	11.4	14.5	9.2 - 19.1	105	17.63	1.06	6.0	17.7	14.4 - 20.9
Sysmex XN-1000	14	15.07	1.17	7.8	14.7	11.5 - 18.6	14	17.76	1.40	7.9	17.2	13.5 - 22.0
Sysmex XN-430	27	14.13	2.21	15.7	14.9	7.4 - 20.8	27	17.87	0.90	5.0	18.1	15.1 - 20.6
Sysmex XN-550	17	13.82	1.86	13.5	14.2	8.2 - 19.4	17	17.20	2.12	12.4	17.8	10.8 - 23.6
Sysmex XS-1000i	36	13.37	1.92	14.4	13.3	7.6 - 19.2	36	17.39	1.01	5.8	17.6	14.3 - 20.5
<u>Instrument</u>	Specimen MX-10											
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>						
All Method	110	14.48	1.79	12.3	14.4	9.1 - 19.9						
All Sysmex XE/XT Instruments	10	17.82	2.85	16.0	18.4	9.2 - 26.4						
All Sysmex XN/XS Instruments	105	14.49	1.47	10.1	14.4	10.0 - 18.9						
Sysmex XN-1000	14	15.21	1.80	11.8	15.2	9.8 - 20.7						
Sysmex XN-430	27	14.76	1.01	6.8	15.0	11.7 - 17.8						
Sysmex XN-550	17	14.08	2.06	14.7	14.5	7.8 - 20.3						
Sysmex XS-1000i	36	13.62	1.91	14.0	13.9	7.9 - 19.4						

HEMATOLOGY W/ 5 or 6-PART DIFFERENTIAL – MONOCYTES (percent)

<i>Instrument</i>	Specimen MX-6						Specimen MX-7					
	<i>Labs</i>	<i>Mean</i>	<i>SD</i>	<i>CV</i>	<i>Median</i>	<i>Range</i>	<i>Labs</i>	<i>Mean</i>	<i>SD</i>	<i>CV</i>	<i>Median</i>	<i>Range</i>
All Method	110	1.45	0.84	58.2	1.2	0.0 - 4.0	110	0.87	0.45	51.4	0.7	0.0 - 2.3
All Sysmex XE/XT Instruments	10	2.22	1.01	45.3	2.0	0.0 - 5.3	10	1.58	0.75	47.2	1.3	0.0 - 3.9
All Sysmex XN/XS Instruments	105	1.41	0.82	58.2	1.2	0.0 - 3.9	105	0.83	0.40	47.7	0.7	0.0 - 2.1
Sysmex XN-1000	14	2.46	0.93	37.8	2.5	0.0 - 5.3	14	1.41	0.56	39.5	1.7	0.0 - 3.1
Sysmex XN-430	27	0.96	0.35	36.7	0.9	0.0 - 2.1	27	0.63	0.13	21.3	0.6	0.2 - 1.1
Sysmex XN-550	17	1.24	0.38	31.0	1.1	0.0 - 2.4	17	0.76	0.24	31.7	0.7	0.0 - 1.5
Sysmex XS-1000i	36	0.94	0.35	37.8	0.8	0.0 - 2.1	36	0.67	0.19	28.9	0.7	0.0 - 1.3
<i>Instrument</i>	Specimen MX-8						Specimen MX-9					
	<i>Labs</i>	<i>Mean</i>	<i>SD</i>	<i>CV</i>	<i>Median</i>	<i>Range</i>	<i>Labs</i>	<i>Mean</i>	<i>SD</i>	<i>CV</i>	<i>Median</i>	<i>Range</i>
All Method	110	0.58	0.38	66.1	0.5	0.0 - 1.8	110	0.88	0.43	48.6	0.8	0.0 - 2.2
All Sysmex XE/XT Instruments	10	1.20	0.48	40.4	1.3	0.0 - 2.7	10	1.76	0.90	51.1	1.7	0.0 - 4.5
All Sysmex XN/XS Instruments	105	0.54	0.35	63.8	0.4	0.0 - 1.6	105	0.86	0.41	47.7	0.7	0.0 - 2.1
Sysmex XN-1000	14	0.84	0.45	53.8	0.9	0.0 - 2.3	14	1.49	0.57	38.7	1.6	0.0 - 3.3
Sysmex XN-430	27	0.30	0.10	34.3	0.3	0.0 - 0.7	27	0.71	0.28	39.3	0.6	0.0 - 1.6
Sysmex XN-550	17	0.62	0.46	74.3	0.3	0.0 - 2.1	17	0.88	0.33	38.0	0.9	0.0 - 1.9
Sysmex XS-1000i	36	0.30	0.12	40.8	0.3	0.0 - 0.7	36	0.77	0.25	31.9	0.7	0.0 - 1.6
<i>Instrument</i>	Specimen MX-10											
	<i>Labs</i>	<i>Mean</i>	<i>SD</i>	<i>CV</i>	<i>Median</i>	<i>Range</i>						
All Method	110	0.55	0.35	64.8	0.5	0.0 - 1.7						
All Sysmex XE/XT Instruments	10	1.44	0.69	48.0	1.3	0.0 - 3.6						
All Sysmex XN/XS Instruments	105	0.50	0.29	57.3	0.5	0.0 - 1.4						
Sysmex XN-1000	14	0.82	0.46	56.3	0.8	0.0 - 2.3						
Sysmex XN-430	27	0.35	0.21	59.0	0.3	0.0 - 1.0						
Sysmex XN-550	17	0.30	0.01	0.0	0.3	0.2 - 0.4						
Sysmex XS-1000i	36	0.31	0.26	83.2	0.3	0.0 - 1.1						

2019 M2 Specimens BC-7 through BC-12

CASE HISTORY:


A 4-year-old female was brought to the emergency department by her parents. The patient had abdominal pain, nausea, vomiting, and loose, watery stools. Upon physical examination, the patient appeared ill, dehydrated, lethargic, and febrile. A CBC was performed, and the results appear below.

TEST	RESULTS	REFERENCE RANGE
WBC	3.0 x 10 ⁹ /L	5 - 17 x 10 ⁹ /L
RBC	3.8 x 10 ¹² /L	4.0 - 5.2 x 10 ¹² /L
HGB	9.9 g/dL	10.2 - 15.2 g/dL
HCT	30 %	34 - 48 %
MCV	79 fL	78 - 94 fL
MCH	26 pg	23 - 31 pg
MCHC	33 g/dL	32 - 36 g/dL
RDW	19 %	11 - 15 %
PLT	85 x 10 ⁹ /L	150 - 450 x 10 ⁹ /L

This patient was diagnosed with acute lymphoblastic leukemia and septicemia. Acute lymphoblastic leukemia (ALL) is the most common cancer in children, and accounts for about 20% of acute leukemias in adults. In ALL, the bone marrow produces too many lymphoblasts. These abnormal cells do not develop into mature lymphocytes or function as normal lymphocytes. Patients with ALL are at risk for blood stream infections due to neutropenia and decreased immunoglobulins. Lymphoblasts crowd the bone marrow, leaving less room for normal blood cells to develop. Decreased neutrophil production increases the risk of infection and fever. Decreased red blood cell production leads to anemia and symptoms including fatigue. Decreased platelet production causes bruising and bleeding. There may be enlargement of the lymph nodes, spleen, and liver. Bone pain is a common symptom often overlooked by providers who attribute it to “growing pains.” Bone pain, joint pain, and headaches may be symptoms of lymphoblast infiltration into those areas. Different subgroups of ALL are based on the age of the patient, cell type, and chromosomal changes. Flow cytometry is used to analyze antigens on the malignant cells. This analysis, called immunophenotyping, helps to determine the patient’s prognosis and course of treatment.

BLOOD CELL IDENTIFICATION


Specimen BC-7



<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Platelet, normal	162	100%	Acceptable

The arrows in this photograph point to **normal platelets**. Normal platelets are much smaller than red blood cells, measuring 2-4 microns. They are round to elliptical in shape. They generally have a central area packed with red to violet colored granules, and a clear blue area that surrounds the granules. To view another photo of normal platelets, see 2018 M3 BC-16.


Specimen BC-8



<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Lymphocyte, reactive	111	68.94%	Acceptable
Immature/abnormal cell – refer	33	20.50%	Acceptable
Monocyte	13	8.07%	

The arrow in this photograph points to a **reactive lymphocyte**. Reactive lymphs come in a wide range of forms, sizes, and shapes. These large, atypical lymphocytes respond to immune stimuli in the body. The nucleus is large and appears elongated or stretched. The cytoplasm of the reactive lymph is light blue, with characteristic peripheral basophilia (darker blue color) at the extreme border. This particular cell also has very distinct radial basophilia, which is sometimes seen in reactive lymphocytes in addition to peripheral basophilia. The cytoplasm of the reactive lymph is indented by adjacent red cells and tends to flow around them. In contrast, monocytes have grayer cytoplasm and lack peripheral basophilia, and monos tend to have pseudopods that appear to push away nearby red cells. The monocyte nucleus is often folded, is lighter-staining than that of the lymph, and has a more open, lacy, chromatin structure. To view another photo of a reactive lymphocyte, see 2018 M2 Specimen BC-9. To view a monocyte, see 2019 M1 Specimen BC-1.


Specimen BC-9



<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Neutrophil-seg//band w/toxic	135	83.33%	Acceptable
Neutrophil-Segmented or band	26	16.05%	

The arrow in this photograph points to a segmented neutrophil with **toxic granulation**. The cytoplasm of cells with toxic granulation contains small, coarse, dark purple granules. Toxic granulation is a sign that the patient has a serious infection or other toxic condition. In contrast, normal neutrophils primarily have fine, smooth, pink granules. Basophils contain much larger purple-black granules, which can cover part of the cell's nucleus and protrude from the cytoplasmic edge. To view another photo of a neutrophil with toxic granulation, see 2017 M3 Specimen BC-15. To view a basophil, see 2018 M3 Specimen BC-14.

Specimen BC-10




<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Erythrocyte, normal	134	82.72%	Acceptable
Hypochromic red cell	15	9.26%	
Macrocyte	9	5.56%	

The arrows in this photograph point to **normal erythrocytes** (red blood cells.) The arrowed cells are approximately the same size as the nucleus of the mature, resting lymphocyte nearby. In a normochromic cell, the area of central pallor takes up about one third of the diameter of the cell. To view another photo of normal red cells, see 2018 M3 Specimen BC-17. To view a photo of hypochromic red cells, see 2018 M1 Specimen BC-5.

BLOOD CELL IDENTIFICATION

Specimen BC-11




<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Immature/abnormal cell – refer	83	51.55%	Acceptable
Blast cell	65	40.37%	Acceptable
Prolymphocyte	8	4.97%	

The arrows in this photograph point to **blast cells**, specifically lymphoblasts. Lymphoblasts are the precursor cells to lymphocytes. Blasts seen in the peripheral blood are always an abnormal finding. The blast stage is the earliest stage of white blood cell development, and is normally found only in the bone marrow. Blast cells have a high ratio of nucleus to cytoplasm, and the cytoplasm is deeply basophilic. The round-oval nucleus of the lymphoblast may have a cleft. Compared to reactive lymphs, blasts have less variation in size and staining characteristics, and far less cytoplasm. To view another photo of a blast cell, see 2016 M3 Specimen BC-18.

BLOOD CELL IDENTIFICATION

Specimen BC-12



<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Bacteria, extracellular	134	82.72%	Not graded – Educational Challenge
Immature/abnormal cell-refer	25	15.43%	
Fungi	3	1.85%	

The arrows in this ungraded educational challenge point to **extracellular bacteria**. These rod-shaped organisms are most likely to be gram negative bacilli of enteric origin. Common pathogens associated with abdominal distress in ALL patients include *Pseudomonas aeruginosa*, *Escherichia coli*, *Klebsiella pneumoniae*, and *Salmonella* species. Although they appear purple in this Wright stained preparation, these bacteria might appear pink if they were Gram stained. Although bacteria and parasites are rarely seen in a peripheral blood smear, it is crucial that laboratorians be on the lookout for organisms in and around the cells, as this is a critical finding. To view a photo of intracellular bacteria, see 2007 M1 Specimen BC-6. To view a photo of intracellular and extracellular fungi, see 2018 M2 Specimen BC-12.

References:

Carr, J.H., Rodak, B.F.: *Clinical Hematology Atlas, 3rd ed.* Saunders, St. Louis, 2009.

Larson, R.A. "Acute Leukemia." *ACP Medicine*. Ed. D. C. Dale. New York: WebMD, Inc., 2004. 2503.

O'Connor O'Connor, B. H.: *A Color Atlas and Instruction Manual of Peripheral Blood Cell Morphology*. Williams & Wilkins, Baltimore MD, 1984.

PDQ® Pediatric Treatment Editorial Board. PDQ Childhood Acute Lymphoblastic Leukemia Treatment. Bethesda, MD: National Cancer Institute. Updated 04/26/2019. Available at: <https://www.cancer.gov/types/leukemia/patient/child-all-treatment-pdq>. Accessed 06/26/2019. [PMID: 26389385]

BLOOD BANK

ABO GROUP

<u>Specimen</u>	<u>Results</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
BB-6	Group B	5	100%	Acceptable
BB-7	Group A	5	100%	Acceptable
BB-8	Group O	5	100%	Acceptable
BB-9	Group A	5	100%	Acceptable
BB-10	Group O	5	100%	Acceptable

RH FACTOR (D TYPE)

<u>Specimen</u>	<u>Results</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
BB-6	Rh Negative	17	100%	Acceptable
BB-7	Rh Negative	16	94.12%	Acceptable
	Rh Positive	1	5.88%	
BB-8	Rh Positive	17	100%	Acceptable
BB-9	Rh Positive	17	100%	Acceptable
BB-10	Rh Positive	17	100%	Acceptable

The specimen BB-7 is graded to 100% referee consensus.

UNEXPECTED ANTIBODY DETECTION

<u>Specimen</u>	<u>Results</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
AB-6	No unexpected antibody detected	5	100%	Acceptable
AB-7	Unexpected antibody detected	5	100%	Acceptable
AB-8	No unexpected antibody detect	5	100%	Acceptable
AB-9	No unexpected antibody detected	5	100%	Acceptable
AB-10	No unexpected antibody detected	5	100%	Acceptable

BLOOD BANK

ANTIBODY IDENTIFICATION

<u>Specimen</u>	<u>Results</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
AB-6	No antibody detected	1	100%	Acceptable
AB-7	Anti-D	1	100%	Acceptable
AB-8	No antibody detected	1	100%	Acceptable
AB-9	No antibody detected	1	100%	Acceptable
AB-10	Anti-K	1	100%	Acceptable

COMPATIBILITY TESTING

<u>Specimen</u>	<u>Results</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
AB-6	Compatible	3	100%	Acceptable
AB-7	Not Compatible	3	100%	Acceptable
AB-8	Compatible	3	100%	Acceptable
AB-9	Compatible	3	100%	Acceptable
AB-10	Compatible	3	100%	Acceptable

ACTIVATED PARTIAL THROMBOPLASTIN (seconds)

<u>Reagent/Instrument</u>	Specimen CG-6						Specimen CG-7					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	13	28.6	5.8	20.1	25	24 - 33	13	24.6	1.7	7.0	24	20 - 29
Dade Actin FSL Sysmex CA-500/600 series	6	24.5	0.5	2.2	25	20 - 29	6	23.3	0.5	2.2	23	19 - 27
Hemoliance SynthASil IL ACL, all models	5	38.0	1.4	3.7	38	32 - 44	5	26.5	0.7	2.7	27	22 - 31
Specimen CG-8						Specimen CG-9						
All Method	13	32.0	5.1	16.0	29	27 - 37	13	44.9	2.7	6.0	44	38 - 52
Dade Actin FSL Sysmex CA-500/600 series	6	28.3	0.5	1.8	28	24 - 33	6	43.8	1.5	3.4	44	37 - 51
Hemoliance SynthASil IL ACL, all models	5	35.0	0.1	0.0	35	29 - 41	5	49.5	0.7	1.4	50	42 - 57
Specimen CG-10												
All Method	13	36.8	2.1	5.7	36	31 - 43						
Dade Actin FSL Sysmex CA-500/600 series	6	36.0	0.9	2.5	36	30 - 42						
Hemoliance SynthASil IL ACL, all models	5	40.5	0.7	1.7	41	34 - 47						

Fibrinogen (mg/dL)

One participant reported Fibrinogen. The vendor assay values on a Sysmex CA-540 for specimens CG-6 through CG-10 are: 350 mg/dL, 220 mg/dL, 105 mg/dL, 197 mg/dL, and 194 mg/dL, respectively.

COAGUCHEK XS PLUS PROTHROMBIN TIME-INTERNATIONAL NORMALIZED RATIO (INR)

<u>Instrument</u>	Specimen XS-6						Specimen XS-7					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	36	3.04	0.12	3.9	3.0	2.4 - 3.7	36	2.11	0.07	3.5	2.1	1.6 - 2.6
All Roche CoaguChek XS Plus Instruments	34	3.03	0.12	3.9	3.0	2.4 - 3.7	34	2.11	0.07	3.5	2.1	1.6 - 2.6
Roche CoaguChek XS Plus - Waived	24	3.03	0.11	3.7	3.0	2.4 - 3.7	24	2.11	0.07	3.4	2.1	1.6 - 2.6
Roche CoaguChek XS Plus	10	3.03	0.15	5.0	3.1	2.4 - 3.7	10	2.11	0.09	4.3	2.1	1.6 - 2.6
	Specimen XS-8						Specimen XS-9					
All Method	24	1.16	0.05	4.3	1.2	0.9 - 1.4	24	2.11	0.12	5.8	2.1	1.6 - 2.6
All Roche CoaguChek XS Plus Instruments	21	1.16	0.05	4.5	1.2	0.9 - 1.4	21	2.10	0.12	5.9	2.1	1.6 - 2.6
Roche CoaguChek XS Plus - Waived	10	1.18	0.04	3.5	1.2	0.9 - 1.5	10	2.08	0.10	4.7	2.1	1.6 - 2.5
Roche CoaguChek XS Plus	9	-	-	-	1.1	0.9 - 1.4	9	-	-	-	2.2	1.6 - 2.6
	Specimen XS-10											
All Method	24	1.14	0.05	4.5	1.1	0.9 - 1.4						
All Roche CoaguChek XS Plus Instruments	21	1.13	0.05	4.4	1.1	0.9 - 1.4						
Roche CoaguChek XS Plus - Waived	10	1.12	0.04	3.7	1.1	0.8 - 1.4						
Roche CoaguChek XS Plus	9	-	-	-	1.2	0.9 - 1.4						

COAGUCHEK XS - INTERNATIONAL NORMALIZED RATIO (INR)

<u>Instrument</u>	Specimen INX-3						Specimen INX-4					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
Roche CoaguChek XS	106	1.17	0.05	4.1	1.2	0.9 - 1.5	107	2.14	0.08	3.9	2.1	1.7 - 2.6

i-Stat PROTHROMBIN TIME (seconds)

<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>Specimen PTI-6</u>				<u>Labs</u>	<u>Mean</u>	<u>Specimen PTI-7</u>			
			<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>			<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
i-Stat Prothrombin Time	10	15.58	0.81	5.2	15.2	13.2 - 18.0	10	28.98	0.77	2.7	28.7	24.6 - 33.4
<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>Specimen PTI-8</u>				<u>Labs</u>	<u>Mean</u>	<u>Specimen PTI-9</u>			
			<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>			<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
i-Stat Prothrombin Time	10	28.10	1.32	4.7	28.2	23.8 - 32.4	10	15.26	0.85	5.6	15.0	12.9 - 17.6
<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>Specimen PTI-10</u>				<u>Labs</u>	<u>Mean</u>	<u>Specimen PTI-10</u>			
			<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>			<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
i-Stat Prothrombin Time	10	28.30	0.88	3.1	28.0	24.0 - 32.6						

i-Stat PROTHROMBIN TIME - INTERNATIONAL NORMALIZED RATIO (INR)

<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>Specimen PTI-6</u>				<u>Labs</u>	<u>Mean</u>	<u>Specimen PTI-7</u>			
			<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>			<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
i-Stat Prothrombin Time	10	1.32	0.04	3.4	1.3	1.0 - 1.6	10	2.52	0.04	1.8	2.5	2.0 - 3.1
<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>Specimen PTI-8</u>				<u>Labs</u>	<u>Mean</u>	<u>Specimen PTI-9</u>			
			<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>			<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
i-Stat Prothrombin Time	10	2.44	0.13	5.5	2.5	1.9 - 3.0	10	1.28	0.08	6.5	1.3	1.0 - 1.6
<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>Specimen PTI-10</u>				<u>Labs</u>	<u>Mean</u>	<u>Specimen PTI-10</u>			
			<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>			<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
i-Stat Prothrombin Time	10	2.46	0.09	3.6	2.4	1.9 - 3.0						

FLUID CELL COUNT – WHITE BLOOD CELL COUNT (µL)

<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>Specimen BF-3</u>				<u>Labs</u>	<u>Mean</u>	<u>Specimen BF-4</u>			
			<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>			<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	5	344.0	63.6	18.5	344	216 - 472	5	5.0	2.8	56.6	5	0 - 11

FLUID CELL COUNT – RED BLOOD CELL COUNT (µL)


<u>Instrument</u>	<u>Labs</u>	<u>Mean</u>	<u>Specimen BF-3</u>				<u>Labs</u>	<u>Mean</u>	<u>Specimen BF-4</u>			
			<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>			<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	5	1163.0	9.9	0.9	1163	1143 - 1183	5	0.0	0.1	0.0	0	0 - 1

2019 M2
FLUID CRYSTAL IDENTIFICATION
Specimens FC-3 and FC-4

Crystals can generally be classified as either optically isotropic or anisotropic. Isotropic solids refract light rays equally in all directions throughout the crystalline structure, regardless of the crystal's orientation to the light source. In contrast, anisotropic crystals interact with light in a manner that is dependent upon the alignment of the crystal. Anisotropic crystals have an internal structure that will cause a ray of light to split into two rays, each traveling in a different direction. A light beam hitting the crystal from one direction or angle will react differently than a beam hitting the crystal at a different angle. This property of splitting light is called **birefringence** or double refraction.

Microscopic examination of synovial fluid for crystals is an important diagnostic test in the evaluation of arthritis. Some crystals can be identified by their shape or morphology alone. Others have similar shapes and need specialized techniques for identification. Using compensated polarized light helps us to identify crystals based on the optical differences described above. The compensator separates the microscope's light rays into slow-moving and fast-moving vibrations or waves. The compensator is marked with an arrow indicating the direction of the slow vibration. The "axis" in the photos below indicates the direction of the slow wave. Color produced by a crystal aligned with the slow-vibration ray of light can be used to identify the crystal. This difference in color is due to the molecular structure inside the crystal, which either allows the light to pass through unchanged, or impedes the light.

Specimen FC-3




<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
CPPD crystals	4	100%	Acceptable

The objects in this photograph are **calcium pyrophosphate dihydrate (CPPD) crystals**. Calcium pyrophosphate crystals cause calcium pyrophosphate disease, a condition commonly called pseudogout. CPPD crystals are usually rhomboidal or rod-shaped, but are occasionally needle-shaped. These crystals demonstrate **positive birefringence**, because they are blue in color when aligned with (parallel to) the compensator filter/axis. Crystals that are not aligned parallel to the axis will be pink-red. The large number of white blood cells in this fluid specimen indicates presence of acute inflammation. To view another photo of CPPD crystals, see 2017 M1 Specimen FC-2.

**2019 M2
FLUID CRYSTAL IDENTIFICATION**

Specimen FC-4



<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
MSU (Monosodium Urate) crystals	4	100%	Acceptable

The objects in this photograph are **monosodium urate (MSU) crystals**. MSU crystals are usually thin and needle-like with pointed ends. They can be either intracellular or extracellular. MSU crystals are associated with gout, which is a common crystal-induced inflammatory arthritis. The crystals form in joints and tissues when the uric acid level is elevated. They cause inflammation and soft tissue damage, resulting in painful swelling, usually in one joint. The base of the big toe is often affected. Gout is caused by either decreased excretion of uric acid into the urine, or increased production of uric acid. There are many factors that contribute to gout, including alcohol use, purine-rich diets, obesity and the metabolic syndrome, and dehydration or use of diuretic agents. Since there are other crystals that can be needle-shaped, examination with a red plate compensator can help with identification. MSU crystals are **negatively birefringent**, meaning the crystals that are lying parallel to (aligned with) the compensator filter axis are yellow, and the crystals lying perpendicular to the filter axis are blue. To view another photo of MSU, see 2018 M3 Specimen FC-5.

REFERENCES:

Abramowitz, M, Davidson, MW. "Optical Birefringence." *Olympus Microscopy Resource Center*. Accessed 7/28/16. Available at: <http://www.olympusmicro.com/primer/lightandcolor/birefringence.html>

Al-Ashkar, F. "Gout and Pseudogout." *Cleveland Clinic Center for Continuing Education*. Accessed 7/28/16. Available at: <http://www.clevelandclinicmeded.com/medicalpubs/diseasemanagement/rheumatology/gout-and-pseudogout/default.htm>

Carr, J.H., Rodak, B.F.: *Clinical Hematology Atlas, 3rd ed.* Saunders, St. Louis, 2009.

"Double Refraction / Optics." *Encyclopedia Britannica Online*. Encyclopedia Britannica, n.d. Web. 18 May 2015. Available at: <http://www.britannica.com/EBchecked/topic/170003/double-refraction>

Mundt, L.A, Shanahan, K.: *Graff's Textbook of Routine Urinalysis and Body Fluids, 2nd ed.* Philadelphia: Lippincott Williams & Wilkins, 2011.

Ringsrud, K. M., Linné, J. J.: *Urinalysis and Body Fluids/ A ColorText and Atlas*. St. Louis, Mosby, 1995.

MICROALBUMIN, DIPSTICK

Specimen UM-2

Participant Results

<u>Method</u>	<u>Labs</u>	<u>Negative</u>	<u>10 mg/L</u>	<u>20 mg/L</u>	<u>30 mg/L</u>	<u>50 mg/L</u>	<u>80 mg/L</u>	<u>100 mg/L</u>	<u>150 mg/L</u>	<u>+ (4 - 8 mg/dL)</u>	<u>++ (>8 mg/dL)</u>
ALL METHODS	31	-	-	-	-	-	26	1	4	-	-
Roche Micral - 1 minute	1	-	-	-	-	-	-	1	-	-	-
Siemens Clinitek Microalbumin	28	-	-	-	-	-	24	-	4	-	-
Siemens Clinitek Status / Status+	1	-	-	-	-	-	1	-	-	-	-
Uriscan Optima	1	-	-	-	-	-	1	-	-	-	-

CREATININE, DIPSTICK

Specimen UM-2

Participant Results

<u>Method</u>	<u>Labs</u>	<u>Negative</u>	<u>10 mg/dL</u>	<u>30 mg/dL</u>	<u>50 mg/dL</u>	<u>100 mg/dL</u>	<u>200 mg/dL</u>	<u>300 mg/dL</u>
ALL METHODS	31	-	1	-	2	16	11	1
Siemens Clinitek Microalbumin	28	-	1	-	2	16	8	1
Siemens Clinitek Status / Status+	1	-	-	-	-	-	1	-
Siemens Multistix Pro	2	-	-	-	-	-	2	-

MICROALBUMIN, QUANTITATIVE

Specimen UM-2

<u>Method</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	58	124.33	7.76	6.2	123.3	87.0 - 161.7
Beckman AU	20	121.47	3.79	3.1	121.7	85.0 - 158.0

CREATININE, URINE (mg/dL)

Specimen UM-2

<u>Method</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	53	67.80	6.08	9.0	68.8	56.2 - 79.4
Beckman AU	21	63.15	4.47	7.1	61.2	52.4 - 73.9

WAIVED HEMATOLOGY–HEMOGLOBIN (g/dL)

<u>Instrument</u>	Specimen HD-6						Specimen HD-7					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	95	13.22	0.44	3.3	13.2	12.2 - 14.2	98	16.78	0.63	3.8	16.7	15.5 - 18.1
All Stanbio Methods	21	13.49	0.59	4.4	13.7	12.3 - 14.7	22	17.26	0.74	4.3	17.4	15.7 - 18.8
Alere (Stanbio) HemoPoint H2	21	13.49	0.59	4.4	13.7	12.3 - 14.7	22	17.26	0.74	4.3	17.4	15.7 - 18.8
HemoCue	72	13.16	0.32	2.4	13.1	12.2 - 14.1	74	16.60	0.46	2.7	16.6	15.4 - 17.8

WAIVED HEMATOLOGY–HEMATOCRIT (percent)

<u>Instrument</u>	Specimen HD-6						Specimen HD-7					
	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	16	36.02	7.12	19.8	39.0	21.7 - 50.3	16	49.46	4.07	8.2	50.3	41.3 - 57.6
Alere (Stanbio) HemoPoint H2	10	38.28	4.83	12.6	41.0	28.6 - 48.0	10	50.02	3.19	6.4	51.0	43.6 - 56.4

KOH SKIN PREPARATION

<u>Specimen</u>	<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
K-3	Yeast/fungal elements present	86	76.79%	Acceptable
	Yeast/fungal elements absent	26	23.21%	

Organism present in specimen K-3: *Aspergillus niger*. The specimen K-3 is graded to 91% referee consensus.

K-4	Yeast/fungal elements present	108	96.43%	Acceptable
	Yeast/fungal elements absent	4	3.57%	

Organism present in specimen K-4: *Tritirachium roseum*.

URINALYSIS DIPSTICK–SPECIFIC GRAVITY

Specimen UA-2

<u>Method</u>	<u>Labs</u>	<u>Mean</u>	<u>SD</u>	<u>CV</u>	<u>Median</u>	<u>Range</u>
All Method	693	1.0203	0.0035	0.3	1.020	1.010 - 1.031
All Roche Methods	13	1.0150	0.0020	0.2	1.015	1.005 - 1.025
All Siemens Methods	441	1.0199	0.0023	0.2	1.020	1.009 - 1.030
Diagnostic Test Group Clarity Urocheck 120	10	1.0275	0.0026	0.3	1.028	1.017 - 1.038
Henry Schein Urispec / Urispec Plus	21	1.0169	0.0034	0.3	1.015	1.006 - 1.027
McKesson 120 Urine Analyzer	23	1.0265	0.0023	0.2	1.025	1.016 - 1.037
Roche Chemstrips	29	1.0155	0.0024	0.2	1.015	1.005 - 1.026
Siemens Clinitek 50	11	1.0186	0.0023	0.2	1.020	1.008 - 1.029
Siemens Clinitek Advantus	14	1.0200	0.0001	0.0	1.020	1.010 - 1.030
Siemens Clinitek Status / Status+	403	1.0199	0.0024	0.2	1.020	1.009 - 1.030
Siemens Reagent Strips	99	1.0204	0.0029	0.3	1.020	1.010 - 1.031

URINALYSIS DIPSTICK-pH

Specimen UA-2

Participant Results

<u>Method</u>	<u>Labs</u>	<u>≤3.5</u>	<u>4.0</u>	<u>4.5</u>	<u>5.0</u>	<u>5.5</u>	<u>6.0</u>	<u>6.5</u>	<u>7.0</u>	<u>7.5</u>	<u>8.0</u>	<u>8.5</u>	<u>≥9.0</u>
ALL METHODS	708	-	-	-	296	300	105	4	2	1	-	-	-
BTNX Rapid Response U120/U500	1	-	-	-	-	-	1	-	-	-	-	-	-
Consult Diagnostics Reagent Strips	1	-	-	-	-	1	-	-	-	-	-	-	-
Consult Diagnostics Urine Analyzer	3	-	-	-	1	-	2	-	-	-	-	-	-
CTMI CT-120 Urine Analyzer	5	-	-	-	-	-	5	-	-	-	-	-	-
Diagnostic Test Group Clarity Urocheck	3	-	-	-	1	2	-	-	-	-	-	-	-
Diagnostic Test Group Clarity Urocheck 120	10	-	-	-	-	5	4	-	1	-	-	-	-
Germaine Laboratories AimStrip	2	-	-	-	1	1	-	-	-	-	-	-	-
Germaine Labs AimStrip Urine Analyzer	2	-	-	-	-	1	1	-	-	-	-	-	-
Henry Schein One Step Plus	1	-	-	-	1	-	-	-	-	-	-	-	-
Henry Schein Urispec / Urispec Plus	21	-	-	-	20	-	1	-	-	-	-	-	-
McKesson 10SG Reagent Strips	4	-	-	-	3	-	1	-	-	-	-	-	-
McKesson 120 Urine Analyzer	23	-	-	-	1	7	15	-	-	-	-	-	-
Medline 120 Urine Analyzer	6	-	-	-	-	3	3	-	-	-	-	-	-
Medline Urinalysis Reagent Strips	3	-	-	-	2	-	1	-	-	-	-	-	-
Moore Medical Urine Reagent Strips	1	-	-	-	1	-	-	-	-	-	-	-	-
NDC Pro Advantage	1	-	-	-	-	-	1	-	-	-	-	-	-
Other Dipstick Method	7	-	-	-	4	-	2	1	-	-	-	-	-
Roche Chemstrips	31	-	-	-	31	-	-	-	-	-	-	-	-
Roche cobas u 411	1	-	-	-	1	-	-	-	-	-	-	-	-
Roche Criterion Analyzer	2	-	-	-	2	-	-	-	-	-	-	-	-
Roche SuperUA/ChemstripUA	1	-	-	-	1	-	-	-	-	-	-	-	-
Roche Urisys	9	-	-	-	8	-	-	1	-	-	-	-	-
Siemens Clinitek 10 / 100	5	-	-	-	3	2	-	-	-	-	-	-	-
Siemens Clinitek 50	11	-	-	-	9	2	-	-	-	-	-	-	-
Siemens Clinitek 500	4	-	-	-	-	3	1	-	-	-	-	-	-
Siemens Clinitek Advantus	15	-	-	-	-	11	4	-	-	-	-	-	-
Siemens Clinitek Status / Status+	408	-	-	-	154	245	7	-	1	1	-	-	-
Siemens Hemacombistix	1	-	-	-	1	-	-	-	-	-	-	-	-
Siemens Multistix Pro	3	-	-	-	1	-	2	-	-	-	-	-	-
Siemens Reagent Strips	107	-	-	-	41	11	53	2	-	-	-	-	-
Uriscan Optima	2	-	-	-	2	-	-	-	-	-	-	-	-
UriScan Reagent Strips	2	-	-	-	2	-	-	-	-	-	-	-	-

URINALYSIS DIPSTICK-PROTEIN QUALITATIVE

Specimen UA-2

<u>Method</u>	<i>Participant Results</i>												
	<u>Labs</u>	<u>Negative</u>	<u>Trace</u>	<u>(1+)</u>	<u>(2+)</u>	<u>(3+)</u>	<u>(4+)</u>	<u>10 - 20</u> <u>mg/dL</u>	<u>30 - 70</u> <u>mg/dL</u>	<u>75</u> <u>mg/dL</u>	<u>100 - 200</u> <u>mg/dL</u>	<u>≥300 - 600</u> <u>mg/dL</u>	<u>>600 or ≥1000</u> <u>mg/dL</u>
ALL METHODS	721	706	8	-	1	-	-	3	-	-	2	1	-
BTNX Rapid Response U120/U500	1	1	-	-	-	-	-	-	-	-	-	-	-
Consult Diagnostics Reagent Strips	1	1	-	-	-	-	-	-	-	-	-	-	-
Consult Diagnostics Urine Analyzer	3	3	-	-	-	-	-	-	-	-	-	-	-
CTMI CT-120 Urine Analyzer	5	5	-	-	-	-	-	-	-	-	-	-	-
Diagnostic Test Group Clarity Urocheck	3	3	-	-	-	-	-	-	-	-	-	-	-
Diagnostic Test Group Clarity Urocheck 120	10	9	-	-	1	-	-	-	-	-	-	-	-
Germaine Laboratories AimStrip	2	1	1	-	-	-	-	-	-	-	-	-	-
Germaine Labs AimStrip Urine Analyzer	2	2	-	-	-	-	-	-	-	-	-	-	-
Henry Schein One Step Plus	1	1	-	-	-	-	-	-	-	-	-	-	-
Henry Schein Urispec / Urispec Plus	21	21	-	-	-	-	-	-	-	-	-	-	-
McKesson 10SG Reagent Strips	4	4	-	-	-	-	-	-	-	-	-	-	-
McKesson 120 Urine Analyzer	23	23	-	-	-	-	-	-	-	-	-	-	-
Medline 120 Urine Analyzer	6	6	-	-	-	-	-	-	-	-	-	-	-
Medline Urinalysis Reagent Strips	3	3	-	-	-	-	-	-	-	-	-	-	-
Moore Medical Urine Reagent Strips	1	1	-	-	-	-	-	-	-	-	-	-	-
NDC Pro Advantage	1	1	-	-	-	-	-	-	-	-	-	-	-
Other Dipstick Method	7	4	-	-	-	-	-	3	-	-	-	-	-
Roche Chemstrips	41	39	2	-	-	-	-	-	-	-	-	-	-
Roche cobas u 411	1	1	-	-	-	-	-	-	-	-	-	-	-

URINALYSIS DIPSTICK-PROTEIN QUALITATIVE (cont'd)

Specimen UA-2

<u>Method</u>	<u>Participant Results</u>												
	<u>Labs</u>	<u>Negative</u>	<u>Trace</u>	<u>(1+)</u>	<u>(2+)</u>	<u>(3+)</u>	<u>(4+)</u>	<u>10 - 20</u> <u>mg/dL</u>	<u>30 - 70</u> <u>mg/dL</u>	<u>75</u> <u>mg/dL</u>	<u>100 - 200</u> <u>mg/dL</u>	<u>≥300 - 600</u> <u>mg/dL</u>	<u>>600 or ≥1000</u> <u>mg/dL</u>
Roche Criterion Analyzer	2	2	-	-	-	-	-	-	-	-	-	-	-
Roche SuperUA/ChemstripUA	1	1	-	-	-	-	-	-	-	-	-	-	-
Roche Urisys	9	9	-	-	-	-	-	-	-	-	-	-	-
Siemens Clinitek 10 / 100	5	5	-	-	-	-	-	-	-	-	-	-	-
Siemens Clinitek 50	11	11	-	-	-	-	-	-	-	-	-	-	-
Siemens Clinitek 500	4	4	-	-	-	-	-	-	-	-	-	-	-
Siemens Clinitek Advantus	15	15	-	-	-	-	-	-	-	-	-	-	-
Siemens Clinitek Status / Status+	408	406	-	-	-	-	-	-	-	1	1	-	-
Siemens Hemacombistix	1	1	-	-	-	-	-	-	-	-	-	-	-
Siemens Multistix Pro	3	3	-	-	-	-	-	-	-	-	-	-	-
Siemens Reagent Strips	109	103	5	-	-	-	-	-	-	1	-	-	-
Siemens Uristix	2	2	-	-	-	-	-	-	-	-	-	-	-
UriScan Optima	2	2	-	-	-	-	-	-	-	-	-	-	-
UriScan Reagent Strips	2	2	-	-	-	-	-	-	-	-	-	-	-

URINALYSIS DIPSTICK–GLUCOSE

Specimen UA-2

<u>Method</u>	<u>Labs</u>	<u>Negative or Normal</u>	<u>Trace</u>	<u>(1+)</u>	<u>Participant Results</u>				<u>30 - 100 mg/dL</u>	<u>150 - 300 mg/dL</u>	<u>500 mg/dL</u>	<u>>500 or ≥1000 or ≥2000 mg/dL</u>
					<u>(2+)</u>	<u>(3+)</u>	<u>(4+)</u>					
ALL METHODS	723	25	54	150	4	4	3	139	273	7	64	
BTNX Rapid Response U120/U500	1	-	-	1	-	-	-	-	-	-	-	
Consult Diagnostics Reagent Strips	1	-	-	1	-	-	-	-	-	-	-	
Consult Diagnostics Urine Analyzer	3	-	-	1	-	-	-	1	1	-	-	
CTMI CT-120 Urine Analyzer	5	-	-	1	-	-	-	1	3	-	-	
Diagnostic Test Group Clarity Urocheck	3	-	-	3	-	-	-	-	-	-	-	
Diagnostic Test Group Clarity Urocheck 120	10	1	-	7	-	-	-	1	1	-	-	
Germaine Laboratories AimStrip	2	-	-	1	-	-	-	1	-	-	-	
Germaine Labs AimStrip Urine Analyzer	2	-	-	1	-	-	-	-	1	-	-	
Henry Schein One Step Plus	1	-	-	-	-	-	-	-	-	-	1	
Henry Schein Urispec / Urispec Plus	21	-	-	-	-	-	-	1	1	3	16	
McKesson 10SG Reagent Strips	4	-	-	3	-	-	-	-	1	-	-	
McKesson 120 Urine Analyzer	23	-	-	20	-	-	-	1	2	-	-	
Medline 120 Urine Analyzer	6	-	-	5	1	-	-	-	-	-	-	
Medline Urinalysis Reagent Strips	3	-	1	2	-	-	-	-	-	-	-	
Moore Medical Urine Reagent Strips	1	-	-	-	1	-	-	-	-	-	-	
NDC Pro Advantage	1	-	-	1	-	-	-	-	-	-	-	
Other Dipstick Method	7	1	2	-	-	-	-	4	-	-	-	
Roche Chemstrips	41	-	-	-	-	-	2	-	1	4	34	
Roche cobas u 411	1	-	-	-	-	-	-	-	-	-	1	
Roche Criterion Analyzer	2	-	-	-	-	-	-	-	-	-	2	
Roche SuperUA/ChemstripUA	1	-	-	-	-	-	-	-	-	-	1	
Roche Urisys	9	-	-	-	-	2	-	-	-	-	7	
Siemens Clinitek 10 / 100	5	1	-	-	-	-	-	2	2	-	-	
Siemens Clinitek 50	11	-	-	5	-	-	-	2	4	-	-	
Siemens Clinitek 500	4	-	-	1	-	-	-	2	1	-	-	
Siemens Clinitek Advantus	15	-	7	1	-	-	-	5	2	-	-	
Siemens Clinitek Status / Status+	407	3	32	83	-	-	-	82	207	-	-	
Siemens Hemacombistix	1	-	-	-	-	-	-	-	1	-	-	
Siemens Multistix Pro	3	1	1	-	-	-	-	-	1	-	-	
Siemens Reagent Strips	110	15	8	9	1	-	1	34	41	-	1	
Siemens Uristix	3	2	1	-	-	-	-	-	-	-	-	
Uriscan Optima	2	-	-	-	-	1	-	-	1	-	-	
UriScan Reagent Strips	2	-	-	-	1	-	-	-	1	-	-	

URINALYSIS DIPSTICK–KETONES

Specimen UA-2

<u>Method</u>	<u>Labs</u>	<u>Participant Results</u>													
		<u>Negative</u>	<u>Trace</u>	<u>Small</u>	<u>Moderate</u>	<u>Large</u>	<u>(1+)</u>	<u>(2+)</u>	<u>(3+)</u>	<u>(4+)</u>	<u>5 - 10</u> <u>mg/dL</u>	<u>15 - 25</u> <u>mg/dL</u>	<u>40 - 60</u> <u>mg/dL</u>	<u>80 - 100</u> <u>mg/dL</u>	<u>≥150</u> <u>mg/dL</u>
ALL METHODS	707	9	-	9	37	31	15	168	27	1	-	29	341	40	-
BTNX Rapid Response U120/U500	1	-	-	-	-	-	-	1	-	-	-	-	-	-	-
Consult Diagnostics Reagent Strips	1	-	-	-	-	-	-	1	-	-	-	-	-	-	-
Consult Diagnostics Urine Analyzer	3	-	-	-	-	-	-	2	-	-	-	-	1	-	-
CTMI CT-120 Urine Analyzer	5	-	-	-	-	-	-	2	-	-	-	2	1	-	-
Diagnostic Test Group Clarity Urocheck	3	-	-	-	-	-	-	1	2	-	-	-	-	-	-
Diagnostic Test Group Clarity Urocheck 120	10	1	-	-	-	-	1	4	3	-	-	-	1	-	-
Germaine Laboratories AimStrip	2	-	-	-	-	-	-	1	-	-	-	-	1	-	-
Germaine Laboratories AimTab	1	-	-	-	1	-	-	-	-	-	-	-	-	-	-
Germaine Labs AimStrip Urine Analyzer	2	-	-	-	-	-	-	1	-	-	-	-	-	1	-
Henry Schein One Step Plus	1	-	-	-	-	-	-	-	-	-	-	1	-	-	-
Henry Schein Urispec / Urispec Plus	21	-	-	-	-	1	-	-	1	-	-	19	-	-	-
McKesson 10SG Reagent Strips	4	-	-	-	-	-	-	2	2	-	-	-	-	-	-
McKesson 120 Urine Analyzer	23	-	-	-	-	-	-	15	4	-	-	-	4	-	-
Medline 120 Urine Analyzer	6	-	-	-	-	-	-	3	3	-	-	-	-	-	-
Medline Urinalysis Reagent Strips	3	-	-	1	-	-	-	1	1	-	-	-	-	-	-
NDC Pro Advantage	1	-	-	-	-	-	-	1	-	-	-	-	-	-	-
Other Dipstick Method	7	-	-	-	3	-	2	1	-	-	-	-	1	-	-
Roche Chemstrips	32	3	-	5	6	4	-	7	6	1	-	-	-	-	-
Roche cobas u 411	1	-	-	-	-	-	-	-	-	-	-	-	1	-	-
Roche Criterion Analyzer	2	-	-	-	-	-	-	-	-	-	-	-	2	-	-
Roche SuperUA/ChemstripUA	1	-	-	-	-	1	-	-	-	-	-	-	-	-	-
Roche Urisys	9	-	-	-	-	-	-	2	-	-	-	1	6	-	-
Siemens Clinitek 10 / 100	5	-	-	-	-	-	-	1	-	-	-	-	4	-	-
Siemens Clinitek 50	11	-	-	-	-	-	-	4	-	-	-	-	7	-	-
Siemens Clinitek 500	4	-	-	-	-	-	-	1	-	-	-	1	2	-	-
Siemens Clinitek Advantus	14	-	-	-	-	-	6	2	-	-	-	1	5	-	-
Siemens Clinitek Status / Status+	407	3	-	1	3	-	4	103	4	-	-	3	280	6	-
Siemens Multistix Pro	3	-	-	-	-	-	-	-	-	-	-	-	-	3	-
Siemens Reagent Strips	108	-	-	2	23	25	-	8	1	-	-	-	19	30	-
Siemens Uristix	1	1	-	-	-	-	-	-	-	-	-	-	-	-	-
Uriscan Optima	2	-	-	-	-	-	-	1	-	-	-	-	1	-	-
UriScan Reagent Strips	1	-	-	-	-	-	-	-	-	-	-	-	1	-	-

URINALYSIS DIPSTICK–BILIRUBIN

Specimen UA-2

<u>Method</u>	<u>Labs</u>	<u>Negative</u>	<u>Trace</u>	<u>Small</u>	<u>Moderate</u>	<u>Participant Results</u>					<u>0.5 - 1.0</u> <u>mg/dL</u>	<u>2.0 - 4.0</u> <u>mg/dL</u>	<u>6.0 - 10.0</u> <u>mg/dL</u>	<u>>10.0</u> <u>mg/dL</u>
						<u>Large</u>	<u>(1+)</u>	<u>(2+)</u>	<u>(3+)</u>	<u>(4+)</u>				
ALL METHODS	690	688	1	1	-	-	-	-	-	-	-	-	-	
BTNX Rapid Response U120/U500	1	1	-	-	-	-	-	-	-	-	-	-	-	
Consult Diagnostics Reagent Strips	1	1	-	-	-	-	-	-	-	-	-	-	-	
Consult Diagnostics Urine Analyzer	3	3	-	-	-	-	-	-	-	-	-	-	-	
CTMI CT-120 Urine Analyzer	5	5	-	-	-	-	-	-	-	-	-	-	-	
Diagnostic Test Group Clarity Urocheck	3	3	-	-	-	-	-	-	-	-	-	-	-	
Diagnostic Test Group Clarity Urocheck 120	9	9	-	-	-	-	-	-	-	-	-	-	-	
Germaine Laboratories AimStrip	2	2	-	-	-	-	-	-	-	-	-	-	-	
Germaine Labs AimStrip Urine Analyzer	2	2	-	-	-	-	-	-	-	-	-	-	-	
Henry Schein One Step Plus	1	1	-	-	-	-	-	-	-	-	-	-	-	
Henry Schein Urispec / Urispec Plus	21	21	-	-	-	-	-	-	-	-	-	-	-	
McKesson 10SG Reagent Strips	4	4	-	-	-	-	-	-	-	-	-	-	-	
McKesson 120 Urine Analyzer	22	22	-	-	-	-	-	-	-	-	-	-	-	
Medline 120 Urine Analyzer	7	7	-	-	-	-	-	-	-	-	-	-	-	
Medline Urinalysis Reagent Strips	3	3	-	-	-	-	-	-	-	-	-	-	-	
Moore Medical Urine Reagent Strips	1	1	-	-	-	-	-	-	-	-	-	-	-	
NDC Pro Advantage	1	1	-	-	-	-	-	-	-	-	-	-	-	
Other Dipstick Method	4	4	-	-	-	-	-	-	-	-	-	-	-	
Roche Chemstrips	31	31	-	-	-	-	-	-	-	-	-	-	-	
Roche cobas u 411	1	1	-	-	-	-	-	-	-	-	-	-	-	
Roche Criterion Analyzer	2	2	-	-	-	-	-	-	-	-	-	-	-	
Roche SuperUA/ChemstripUA	1	1	-	-	-	-	-	-	-	-	-	-	-	
Roche Urisys	9	9	-	-	-	-	-	-	-	-	-	-	-	
Siemens Clinitek 10 / 100	5	5	-	-	-	-	-	-	-	-	-	-	-	
Siemens Clinitek 50	11	11	-	-	-	-	-	-	-	-	-	-	-	
Siemens Clinitek 500	4	4	-	-	-	-	-	-	-	-	-	-	-	
Siemens Clinitek Advantus	14	14	-	-	-	-	-	-	-	-	-	-	-	
Siemens Clinitek Status / Status+	403	403	-	-	-	-	-	-	-	-	-	-	-	
Siemens Multistix Pro	3	2	-	1	-	-	-	-	-	-	-	-	-	
Siemens Reagent Strips	102	102	-	-	-	-	-	-	-	-	-	-	-	
Uriscan Optima	2	2	-	-	-	-	-	-	-	-	-	-	-	
UriScan Reagent Strips	1	1	-	-	-	-	-	-	-	-	-	-	-	

URINALYSIS DIPSTICK–UROBILINOGEN

Specimen UA-2

<u>Method</u>	<u>Labs</u>	<i>Participant Results</i>				
		<u>Normal or 0.0 - 0.2 mg/dL or <3.2 µmol/L</u>	<u>1.0 or <2.0 mg/dL or 16 or 17 µmol/L</u>	<u>2.0/3.0 mg/dL or 34 or 35 µmol/L</u>	<u>4.0 or 4.0/6.0 mg/dL or 70 µmol/L</u>	<u>≥8.0 or ≥12.0 mg/dL or ≥140 or 200 µmol/L</u>
ALL METHODS	684	681	-	3	-	-
BTNX Rapid Response U120/U500	1	1	-	-	-	-
Consult Diagnostics Reagent Strips	2	2	-	-	-	-
Consult Diagnostics Urine Analyzer	3	3	-	-	-	-
CTMI CT-120 Urine Analyzer	5	5	-	-	-	-
Diagnostic Test Group Clarity Urocheck	3	3	-	-	-	-
Diagnostic Test Group Clarity Urocheck 120	10	10	-	-	-	-
Germaine Laboratories AimStrip	2	2	-	-	-	-
Germaine Labs AimStrip Urine Analyzer	2	2	-	-	-	-
Henry Schein One Step Plus	1	1	-	-	-	-
Henry Schein Urispec / Urispec Plus	22	22	-	-	-	-
McKesson 10SG Reagent Strips	4	4	-	-	-	-
McKesson 120 Urine Analyzer	22	22	-	-	-	-
Medline 120 Urine Analyzer	6	4	-	2	-	-
Medline Urinalysis Reagent Strips	3	3	-	-	-	-
Moore Medical Urine Reagent Strips	1	1	-	-	-	-
NDC Pro Advantage	1	1	-	-	-	-
Other Dipstick Method	4	4	-	-	-	-
Roche Chemstrips	29	29	-	-	-	-
Roche cobas u 411	1	1	-	-	-	-
Roche Criterion Analyzer	2	2	-	-	-	-
Roche SuperUA/ChemstripUA	1	1	-	-	-	-
Roche Urisys	10	10	-	-	-	-
Siemens Clinitek 10 / 100	5	5	-	-	-	-
Siemens Clinitek 50	12	12	-	-	-	-
Siemens Clinitek 500	4	4	-	-	-	-
Siemens Clinitek Advantus	13	13	-	-	-	-
Siemens Clinitek Status / Status+	398	397	-	1	-	-
Siemens Multistix Pro	3	3	-	-	-	-
Siemens Reagent Strips	104	104	-	-	-	-
Uriscan Optima	2	2	-	-	-	-
UriScan Reagent Strips	1	1	-	-	-	-

URINALYSIS DIPSTICK–BLOOD/HEMOGLOBIN

Specimen UA-2

<u>Method</u>	<u>Participant Results</u>											<u>5 - 25</u> <u>Erv/µL</u>	<u>50 -</u> <u>100</u> <u>Erv/µL</u>	<u>200 -</u> <u>250</u> <u>Erv/µL</u>	<u>±0.03</u> <u>mg/dL</u>	<u>0.06</u> : <u>0.10</u> <u>mg/</u> <u>dL</u>	<u>0.2 -</u> <u>0.5</u> <u>mg/</u> <u>dL</u>	<u>≥ 1.0</u> <u>mg/</u> <u>dL</u>
	<u>Labs</u>	<u>Negative</u>	<u>Trace</u>	<u>Small</u>	<u>Moderate</u>	<u>Large</u>	<u>(1+)</u>	<u>(2+)</u>	<u>(3+)</u>	<u>(4+)</u>	<u>(5+)</u>							
ALL METHODS	715	700	9	-	1	2	2	-	1	-	-	-	-	-	-	-	-	
BTNX Rapid Response U120/U500	1	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Consult Diagnostics Reagent Strips	2	2	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Consult Diagnostics Urine Analyzer	3	3	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
CTMI CT-120 Urine Analyzer	5	5	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Diagnostic Test Group Clarity Urocheck	3	3	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Diagnostic Test Group Clarity Urocheck 120	9	8	-	-	-	-	-	-	1	-	-	-	-	-	-	-	-	
Germaine Laboratories AimStrip	2	2	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Germaine Labs AimStrip Urine Analyzer	2	2	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Henry Schein One Step Plus	1	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Henry Schein Urispec / Urispec Plus	21	20	-	-	-	-	1	-	-	-	-	-	-	-	-	-	-	
McKesson 10SG Reagent Strips	4	4	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
McKesson 120 Urine Analyzer	23	23	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Medline 120 Urine Analyzer	6	6	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Medline Urinalysis Reagent Strips	3	3	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Moore Medical Urine Reagent Strips	1	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
NDC Pro Advantage	1	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Other Dipstick Method	7	7	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Roche Chemstrips	39	39	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Roche cobas u 411	1	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Roche Criterion Analyzer	2	2	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Roche SuperUA/ChemstripUA	1	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Roche Urisys	10	10	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Siemens Clinitek 10 / 100	5	4	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Siemens Clinitek 50	12	12	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Siemens Clinitek 500	4	4	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Siemens Clinitek Advantus	15	15	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Siemens Clinitek Status / Status+	405	393	8	-	1	2	1	-	-	-	-	-	-	-	-	-	-	
Siemens Hemacombistix	1	1	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Siemens Multistix Pro	3	3	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Siemens Reagent Strips	108	108	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
Uriscan Optima	2	2	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	
UriScan Reagent Strips	2	2	-	-	-	-	-	-	-	-	-	-	-	-	-	-	-	

URINALYSIS DIPSTICK--LEUKOCYTE ESTERASE

Specimen UA-2

Participant Results

<u>Method</u>	<u>Labs</u>	<u>Negative</u>	<u>Trace</u>	<u>Small</u>	<u>Moderate</u>	<u>Large</u>	<u>(1+)</u>	<u>(2+)</u>	<u>(3+)</u>	<u>(4+)</u>	<u>15 or 25 µL</u>	<u>75 or 100 µL</u>	<u>250 or 500 µL</u>
ALL METHODS	717	10	18	181	191	13	98	152	15	-	3	2	34
BTNX Rapid Response U120/U500	1	-	-	-	-	-	-	1	-	-	-	-	-
Consult Diagnostics Reagent Strips	2	-	1	-	-	-	1	-	-	-	-	-	-
Consult Diagnostics Urine Analyzer	3	-	-	-	-	-	-	2	-	-	-	1	-
CTMI CT-120 Urine Analyzer	5	1	-	-	-	-	-	1	-	-	3	-	-
Diagnostic Test Group Clarity Urocheck	3	-	-	-	-	-	-	3	-	-	-	-	-
Diagnostic Test Group Clarity Urocheck 120	10	1	-	-	-	-	2	4	2	-	-	-	1
Germaine Laboratories AimStrip	2	-	-	-	-	-	1	1	-	-	-	-	-
Germaine Labs AimStrip Urine Analyzer	2	-	-	-	-	-	1	1	-	-	-	-	-
Henry Schein One Step Plus	1	-	-	-	-	-	-	-	-	-	-	-	1
Henry Schein Urispec / Urispec Plus	20	-	-	1	-	-	-	-	1	-	-	-	18
McKesson 10SG Reagent Strips	4	-	-	-	-	-	1	3	-	-	-	-	-
McKesson 120 Urine Analyzer	23	-	-	-	-	-	7	12	3	-	-	1	-
Medline 120 Urine Analyzer	6	-	-	-	-	-	1	4	1	-	-	-	-
Medline Urinalysis Reagent Strips	3	1	-	-	-	-	1	1	-	-	-	-	-
Moore Medical Urine Reagent Strips	1	-	-	-	-	-	-	1	-	-	-	-	-
NDC Pro Advantage	1	-	-	-	-	-	1	-	-	-	-	-	-
Other Dipstick Method	7	1	-	-	2	1	3	-	-	-	-	-	-
Roche Chemstrips	39	-	-	-	-	5	-	34	-	-	-	-	-
Roche cobas u 411	1	-	-	-	-	-	-	-	-	-	-	-	1
Roche Criterion Analyzer	2	-	-	-	-	-	-	-	-	-	-	-	2
Roche SuperUA/ChemstripUA	1	-	-	-	-	-	-	1	-	-	-	-	-
Roche Urisys	10	-	-	-	-	-	-	2	-	-	-	-	8
Siemens Clinitek 10 / 100	5	-	-	4	-	-	-	1	-	-	-	-	-
Siemens Clinitek 50	12	-	-	6	1	-	4	1	-	-	-	-	-
Siemens Clinitek 500	4	-	1	-	2	-	1	-	-	-	-	-	-
Siemens Clinitek Advantus	15	-	-	5	2	-	5	3	-	-	-	-	-
Siemens Clinitek Status / Status+	406	4	12	143	134	5	55	48	5	-	-	-	-
Siemens Multistix Pro	3	-	-	-	3	-	-	-	-	-	-	-	-
Siemens Reagent Strips	109	2	3	20	46	1	12	25	-	-	-	-	-
Siemens Uristix	1	-	-	1	-	-	-	-	-	-	-	-	-
Uriscan Optima	2	-	-	-	-	-	-	-	1	-	-	-	1
UriScan Reagent Strips	2	-	-	-	-	-	-	-	1	-	-	-	1

URINALYSIS DIPSTICK–NITRITE**Specimen UA-2*****Participant Results***

<u>Method</u>	<u>Labs</u>	<u>Negative</u>	<u>Positive</u>
ALL METHODS	716	7	709
BTNX Rapid Response U120/U500	1	-	1
Consult Diagnostics Reagent Strips	2	-	2
Consult Diagnostics Urine Analyzer	3	-	3
CTMI CT-120 Urine Analyzer	5	-	5
Diagnostic Test Group Clarity Urocheck	3	-	3
Diagnostic Test Group Clarity Urocheck 120	10	1	9
Germaine Laboratories AimStrip	2	-	2
Germaine Labs AimStrip Urine Analyzer	2	-	2
Henry Schein One Step Plus	2	-	2
Henry Schein Urispec / Urispec Plus	20	-	20
McKesson 10SG Reagent Strips	4	-	4
McKesson 120 Urine Analyzer	22	-	22
Medline 120 Urine Analyzer	7	-	7
Medline Urinalysis Reagent Strips	3	-	3
Moore Medical Urine Reagent Strips	1	-	1
NDC Pro Advantage	1	-	1
Other Dipstick Method	7	-	7
Roche Chemstrips	39	-	39
Roche cobas u 411	1	-	1
Roche Criterion Analyzer	2	-	2
Roche SuperUA/ChemstripUA	1	-	1
Roche Urisys	10	-	10
Siemens Clinitek 10 / 100	5	-	5
Siemens Clinitek 50	11	-	11
Siemens Clinitek 500	4	-	4
Siemens Clinitek Advantus	15	-	15
Siemens Clinitek Status / Status+	403	4	399
Siemens Multistix Pro	3	1	2
Siemens Reagent Strips	109	1	108
Siemens Uristix	1	-	1
Uriscan Optima	2	-	2
UriScan Reagent Strips	2	-	2

URINALYSIS –MICROALBUMIN (dipstick only)

Specimen UA-2

<u>Method</u>	<u>Labs</u>	<i>Participant Results</i>									
		<u>Negative</u>	<u>10 mg/L</u>	<u>20 mg/L</u>	<u>30 mg/L</u>	<u>50 mg/L</u>	<u>80 mg/L</u>	<u>100 mg/L</u>	<u>150 mg/L</u>	<u>+ (4 - 8 mg/dL)</u>	<u>++ (>8 mg/dL)</u>
ALL METHODS	57	6	43	4	-	-	3	-	1	-	-
Consult Diagnostics Reagent Strips	1	-	1	-	-	-	-	-	-	-	-
Roche Chemstrips	1	1	-	-	-	-	-	-	-	-	-
Roche Micral - 1 minute	7	4	-	3	-	-	-	-	-	-	-
Siemens Clinitek Microalbumin	38	-	36	-	-	-	1	-	1	-	-
Siemens Clinitek Status / Status+	5	-	4	-	-	-	1	-	-	-	-
Siemens Reagent Strips	1	1	-	-	-	-	-	-	-	-	-

URINALYSIS –URINE hCG**Specimen UA-2**

<u>Method</u>	<i>Participant Results</i>		
	<u>Labs</u>	<u>Positive</u>	<u>Negative</u>
ALL METHODS	412	7	405
Alere Acceava hCG-Urine	2	-	2
Alere Clearview 25 hCG Combo	1	-	1
Alere Clearview hCG Cassette	3	-	3
Alere Clearview hCG Combo II	1	-	1
Alere hCG Cassette	4	-	4
Alfa Scientific Instant View	6	-	6
Beckman Coulter ICON 20 hCG	3	-	3
Beckman Coulter ICON 25 hCG	20	-	20
Beckman Coulter ICON II	4	-	4
BioSign hCG	1	-	1
BTNX Rapid Response hCG	1	-	1
Cardinal Health SP Brand combo	27	1	26
Cardinal Hlth SPBrand-cassette	4	-	4
Clarity Diagnostics hCG Combo	2	-	2
Clarity Diagnostics hCG strip/cassette	11	1	10
CONSULT diagnostics hCG Cassette	52	1	51
CONSULT diagnostics hCG Combo	10	-	10
CONSULT diagnostics hCG Dipstick	29	1	28
Consult Diagnostics Reagent Strips	1	-	1
Germaine Laboratories AimStep Pregnancy	2	-	2
Henry Schein One Step	61	1	60
Henry Schein One Step Plus	1	-	1
Immunostics hCG Detector-urine	1	-	1
Jant Pharmacal Accutest	1	-	1
McKesson hCG Combo Cassette	3	-	3
McKesson hCG Urine Cassette	7	1	6
MediChoice hCG Combi Cassette	2	-	2
MediChoice hCG Urine Cassette	1	-	1
Medline hCG Combo Test Cassette	9	1	8
Medline hCG Test Cassette	4	-	4
Moore Medical hCG Urine	1	-	1
NDC Pro Advantage	1	-	1
PEP (Lab Supply) HCG	1	-	1
Polymedco Poly stat hCG	1	-	1
Quidel QuickVue One-Step Combo	17	-	17
Quidel QuickVue One-Step Urine	35	-	35
Quidel QuickVue+ One-Step Combo	24	-	24
Quidel Sofia hCG	2	-	2

URINALYSIS –URINE hCG (cont'd)

Specimen UA-2

<u>Method</u>	<i>Participant Results</i>		
	<u>Labs</u>	<u>Positive</u>	<u>Negative</u>
Sekisui OSOM - Urine Test	1	-	1
Sekisui OSOM Card Pregnancy	6	-	6
Sekisui OSOM hCG Combo Test	2	-	2
Siemens Clinitek Status / Status+	10	-	10
Stanbio QuPID	8	-	8
Stanbio QuPID Plus	2	-	2
Stanbio TRUE hCG	7	-	7
Sure-Vue hCG - 25mIU	2	-	2
Sure-Vue hCG-STAT	8	-	8

FECAL OCCULT BLOOD

<u>Method</u>	Specimen OC-3			Specimen OC-4		
	<u>Labs</u>	<u>Positive</u>	<u>Negative</u>	<u>Labs</u>	<u>Positive</u>	<u>Negative</u>
ALL METHODS	271	268	3	271	4	267
Alere Clearview iFOBT Complete	1	1	-	1	-	1
Beckman Coulter Hemoccult ICT	37	37	-	37	-	37
Guaiaac (slide) Test	157	156	1	157	1	156
Hemosure iFOB	30	29	1	30	2	28
Other Immunochemical FOB kit	26	25	1	26	1	25
Polymedco OC Auto Micro 80	4	4	-	4	-	4
Polymedco OC-Light iFOB	8	8	-	8	-	8
Quidel QuickVue iFOB	5	5	-	5	-	5

2019 M2
Urine Sediment Identification
SPECIMENS US-3 AND US-4

CASE HISTORY:

A pregnant 38-year-old female presented for her first prenatal visit. Symptoms included mild fatigue and morning sickness. The patient appeared to be in good health, and the history and physical examination were unremarkable. Several routine screening tests were ordered. A urinalysis was performed, and the results appear below.

Color = Yellow
Appearance = Cloudy


Dipstick results:

Specific gravity = 1.030
pH = 7.5
Protein = Negative
Glucose = Negative
Ketones = Negative
Bilirubin = Negative
Urobilinogen = Normal
Blood = Negative
Leukocyte Esterase = Small (1+)
Nitrite = Positive

This patient was diagnosed with an asymptomatic urinary tract infection (UTI). Asymptomatic bacteriuria (ASB) is a condition in which there is a significant number of bacteria present in the urine, with or without pyuria (leukocytes), but the patient is not experiencing any symptoms of infection. It happens in about 15% of pregnancies, and can progress to symptomatic cystitis or pyelonephritis. Pregnant women are more susceptible to UTI due to physical and hormonal changes. Screening urinalysis and culture are performed early in the pregnancy, and bacteriuria is treated with antibiotics to prevent complications such as preterm labor or premature rupture of the membranes. ASB is common in healthy women and in adults or children with urologic abnormalities associated with impaired voiding. However, screening and treatment of asymptomatic bacteriuria are generally not recommended unless the patient is pregnant or about to have an endoscopic urologic surgical procedure.

Urine Sediment Identification

Specimen US-3



<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Bacteria	415	98.14%	Acceptable

The arrows in this photograph point to **bacteria**. Most urinary tract infections are caused by rod-shaped enteric bacilli such as *E. coli*, *Proteus*, *Enterobacter*, and *Klebsiella*. These pathogens make enzymes that convert nitrate, a urine waste product, into nitrite. A positive dipstick nitrite test is a strong indicator of the presence of a significant number of nitrate-converting bacteria. However, a negative result for nitrite does not rule out the possibility of a urinary tract infection. Not all pathogens convert nitrate to nitrite. In addition, urine must stay in the bladder at least 4 hours for the chemical conversion to take place. *E. coli*, a nitrite reducer, causes up to 90% of UTIs, however the nitrite test is positive in only 25% of patients with UTI. Frequent urination is a common symptom of UTI, and can cause a false negative nitrite result because the urine has not remained in the bladder long enough for nitrite to form. This is one of the reasons that a first morning specimen is preferred for urinalysis testing. To view another photo of rod-shaped bacteria, see 2016 M3 Specimen US-5. To view a photo of round cocci bacteria, see 2012 M3 Specimen US-6.

Specimen US-4



<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
White blood cell (WBC)	406	96.44%	Acceptable

The arrows in this photograph point to **white blood cells**. White blood cells (leukocytes) appear granular due to the nuclear material inside. Red blood cells, like the one near the center of the field, generally appear hollow or empty and are smaller in size than WBCs. White cells are increased in patients with infection of the bladder or kidneys. The white blood cells that respond to infection and inflammation produce enzymes called esterases. When many WBCs are present in the urine, the dipstick turns positive for leukocyte esterase, indicating pyuria – urine containing white blood cells or pus. To view another photo of WBCs, see 2017 M1 Specimen US-1.

REFERENCES:

Friel, L.A. *Urinary Tract Infection in Pregnancy*. The Merck Manual Professional Version. Merck Sharp & Dohme Corp., Kenilworth, N.J. Updated Feb. 2019. Available at: <https://www.merckmanuals.com/professional/gynecology-and-obstetrics/pregnancy-complicated-by-disease/urinary-tract-infection-in-pregnancy?query=Asymptomatic%20Bacteriuria>. Accessed June 26, 2019.

Mundt, L.A, Shanahan, K.: *Graff's Textbook of Routine Urinalysis and Body Fluids, 2nd ed*. Philadelphia: Lippincott Williams & Wilkins, 2011.


Nicolle, Lindsay E. "Clinical Practice Guideline for the Management of Asymptomatic Bacteriuria: 2019 Update by the Infectious Diseases Society of America." *Clinical Infectious Diseases*, March 21, 2019. doi:10.1093/cid/ciy1121. Available at: <https://www.idsociety.org/globalassets/idsa/practice-guidelines/2019-asymptomatic-bacteriuria.pdf>

Ringsrud, K. M., Linné, J. J.: *Urinalysis and Body Fluids/ A ColorText and Atlas*. St. Louis: Mosby, 1995.

PROVIDER-PERFORMED MICROSCOPY (PPM)

Wet Mount Preparation

Specimen PPM-7




<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Clue cell	332	68.88%	Acceptable
Squamous epithelial cell	145	30.08%	

The arrows in this photograph of a vaginal wet mount point to **clue cells**. Clue cells are epithelial cells covered with bacteria so the cell appears speckled or glittery. Note that the outer edges of the cell are obscured by the bacteria and appear torn or ragged. It is important to differentiate between normal squamous cells and clue cells because squamous epithelial cells are a normal finding, while clue cells are indicative of bacterial vaginosis (BV). To view another clue cell, see 2015 M1 Specimen PPM-1. To view a normal squamous epithelial cell, see 2016 M1 Specimen PPM-1. The specimen PPM-7 is graded to 82% referee consensus

Scabies Detection

Specimen PPM-8



<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Scabies present	189	98.95%	Acceptable
Scabies absent	2	1.05%	

A scabies mite is present in this photograph of a skin scrapings preparation. The mites burrow into the skin and produce an itchy pimple-like (papular) rash. The diagnosis of scabies is often made only by the patient history and examination of the skin. Identification of the mite, its burrows, eggs, or feces (called scybala) confirms the clinical suspicion of scabies. Scabies can be difficult to find by laboratory testing, though, because mites are often few in number. The scabies mite, or human itch mite (*Sarcoptes scabiei*), is usually spread by prolonged direct personal contact with an infested person. It can also be spread indirectly by prolonged contact with infested clothing or bedding. There may be only 10-15 mites on the entire body of an infested person who is otherwise healthy, so a negative test does not rule out the diagnosis. To view another photo of a scabies mite, see 2017 M1 Specimen PPM-3.

PROVIDER-PERFORMED MICROSCOPY (PPM)

SPERM DETECTION

Specimen PPM-9




<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Sperm absent	288	99.65%	Acceptable
Sperm present	1	0.35%	

Spermatozoa are absent in this photograph of a vaginal wet mount preparation. To view a photo of spermatozoa, see 2017 M3 Specimen PPM-15.

NASAL SMEAR

Specimen PPM-10




<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Eosinophils absent	73	68.22%	Acceptable
Eosinophils present	34	31.78%	

Eosinophils are absent in this photograph of Wright-stained nasal mucus. The cells shown in this photo are not eosinophils. The purpose of examining respiratory secretions for leukocytes (white blood cells) is to differentiate allergic conditions from infections. Eosinophils are a specific type of leukocyte associated with allergic conditions. “Eos” take on a unique red-orange color that makes them easy to spot and identify. The orange color of the eosinophil comes from the dye eosin, which is a component of Wright stain. The cytoplasm of an eosinophil is filled with large, round, orange-staining granules that surround the purple nucleus but do not obscure it from view. To view a photo of eosinophils in a nasal smear, see 2017 M2 Specimen PPM-10. The specimen PPM-10 is graded to 100% referee consensus

PROVIDER-PERFORMED MICROSCOPY (PPM)

STOOL PREPARATION

Specimen PPM-11




<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Leukocytes present	166	98.22%	Acceptable
Leukocytes absent	3	1.78%	

Many leukocytes are present in this photograph of a Wright stained stool preparation. To view another photo of fecal leukocytes, see 2017 M2 Specimen PPM-11.

PINWORM PREPARATION

Specimen PPM-12



<u>Identification</u>	<u>Labs</u>	<u>Percent</u>	<u>Performance</u>
Pinworms/eggs absent	200	97.56%	Acceptable
Pinworms/eggs present	5	2.44%	

Pinworms/eggs are absent in this photograph of a perianal pinworm preparation. To view a photo of pinworm eggs, see 2017 M2 Specimen PPM-12.

REFERENCES:

Fischer, P. M.: *The Office Laboratory*. Norwalk, Conn.: Appleton-Century-Crofts, 1983.

Mundt, L.A, Shanahan, K.: *Graff's Textbook of Routine Urinalysis and Body Fluids, 2nd ed.* Philadelphia: Lippincott Williams & Wilkins, 2011.

Ringsrud, K. M., Linné, J. J.: *Urinalysis and Body Fluids/A ColorText and Atlas*. St. Louis: Mosby, 1995.

Scabies. In: Parasite Image Library at DPDx Laboratory Identification of Parasites of Public Health Concern. Centers for Disease Control and Prevention (CDC). Available at: <http://www.cdc.gov/dpdx/scabies/index.html>

Medical Laboratory Evaluation
25 Massachusetts Ave NW Ste 700
Washington, DC 20001-7401
800-338-2746 • 202-261-4500 • Fax: 202-835-0440